

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2879

Instrument ID: P7

Start Date : 08/20/2025

Run Number: LB136893

End Date : 08/20/2025

| Lab<br>SampleID | Client<br>SampleID | Time | Internal Standard %RI For: Non-Collision Cell |   |                 |   |                |   |                  |   |                  |   |
|-----------------|--------------------|------|-----------------------------------------------|---|-----------------|---|----------------|---|------------------|---|------------------|---|
|                 |                    |      | Element<br>6Li                                | Q | Element<br>45Sc | Q | Element<br>89Y | Q | Element<br>103Rh | Q | Element<br>159Tb | Q |
| S0              | S0                 | 1049 | 100                                           |   | 100             |   | 100            |   | 100              |   | 100              |   |
| S2              | S2                 | 1056 | 102                                           |   | 99              |   | 99             |   | 99               |   | 100              |   |
| S3              | S3                 | 1059 | 103                                           |   | 98              |   | 98             |   | 97               |   | 98               |   |
| S4              | S4                 | 1102 | 105                                           |   | 99              |   | 100            |   | 99               |   | 101              |   |
| S5              | S5                 | 1105 | 104                                           |   | 97              |   | 98             |   | 98               |   | 99               |   |
| S6              | S6                 | 1108 | 106                                           |   | 99              |   | 100            |   | 99               |   | 103              |   |
| S7              | S7                 | 1111 | 107                                           |   | 103             |   | 105            |   | 102              |   | 106              |   |
| S8              | S8                 | 1114 | 104                                           |   | 103             |   | 102            |   | 96               |   | 102              |   |
| ICV01           | ICV01              | 1135 | 95                                            |   | 101             |   | 102            |   | 102              |   | 100              |   |
| LLICV01         | LLICV01            | 1150 | 96                                            |   | 99              |   | 101            |   | 100              |   | 99               |   |
| ICB01           | ICB01              | 1156 | 95                                            |   | 99              |   | 100            |   | 100              |   | 97               |   |
| ICSA01          | ICSA01             | 1200 | 97                                            |   | 98              |   | 101            |   | 98               |   | 100              |   |
| ICSAB01         | ICSAB01            | 1203 | 95                                            |   | 103             |   | 107            |   | 104              |   | 106              |   |
| CCV01           | CCV01              | 1206 | 98                                            |   | 106             |   | 110            |   | 105              |   | 109              |   |
| CCB01           | CCB01              | 1209 | 92                                            |   | 102             |   | 106            |   | 105              |   | 102              |   |
| CRI             | CRI                | 1215 | 93                                            |   | 102             |   | 104            |   | 104              |   | 101              |   |
| PB169300BL      | PB169300BL         | 1219 | 91                                            |   | 101             |   | 103            |   | 102              |   | 99               |   |
| PB169300BS      | PB169300BS         | 1222 | 95                                            |   | 102             |   | 106            |   | 103              |   | 104              |   |
| PB169300TB      | PB169300TB         | 1225 | 91                                            |   | 99              |   | 102            |   | 101              |   | 99               |   |
| Q2879-02        | OU4-TS-GRIL        | 1228 | 93                                            |   | 104             |   | 115            |   | 105              |   | 104              |   |
| Q2879-04        | OU4-TS-GRIL        | 1232 | 89                                            |   | 103             |   | 113            |   | 105              |   | 103              |   |
| Q2879-06        | OU4-TS-GRIL        | 1235 | 87                                            |   | 102             |   | 113            |   | 106              |   | 103              |   |
| Q2879-08        | OU4-TS-GRIL        | 1238 | 87                                            |   | 103             |   | 115            |   | 106              |   | 103              |   |
| Q2879-10        | OU4-TS-GRIL        | 1241 | 87                                            |   | 103             |   | 111            |   | 107              |   | 104              |   |
| Q2879-12        | OU4-TS-GRIL        | 1244 | 87                                            |   | 104             |   | 117            |   | 109              |   | 107              |   |
| CCV02           | CCV02              | 1247 | 92                                            |   | 106             |   | 110            |   | 105              |   | 106              |   |
| CCB02           | CCB02              | 1252 | 90                                            |   | 100             |   | 103            |   | 104              |   | 101              |   |
| Q2879-14        | OU4-TS-GRIL        | 1256 | 92                                            |   | 103             |   | 114            |   | 106              |   | 105              |   |
| Q2879-16        | OU4-TS-GRIL        | 1259 | 89                                            |   | 104             |   | 117            |   | 107              |   | 105              |   |
| Q2879-18        | OU4-TS-GRIL        | 1302 | 88                                            |   | 104             |   | 117            |   | 106              |   | 103              |   |
| Q2879-18DUP     | OU4-TS-GRIL        | 1305 | 88                                            |   | 103             |   | 116            |   | 106              |   | 105              |   |
| Q2879-18L       | OU4-TS-GRIL        | 1308 | 87                                            |   | 104             |   | 109            |   | 106              |   | 103              |   |
| Q2879-18MS      | OU4-TS-GRIL        | 1311 | 89                                            |   | 105             |   | 117            |   | 109              |   | 106              |   |
| Q2879-18MSD     | OU4-TS-GRIL        | 1314 | 88                                            |   | 103             |   | 114            |   | 106              |   | 106              |   |
| Q2879-18A       | OU4-TS-GRIL        | 1317 | 88                                            |   | 106             |   | 117            |   | 108              |   | 105              |   |
| CCV03           | CCV03              | 1327 | 91                                            |   | 108             |   | 112            |   | 106              |   | 107              |   |

Internal Standard %RI Limit: 30 - 120

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2879

Instrument ID: P7

Start Date : 08/20/2025

Run Number: LB136893

End Date : 08/20/2025

| Lab<br>SampleID | Client<br>SampleID | Time | Internal Standard %RI For: Non-Collision Cell |   |                 |   |                |   |                  |   |                  |   |
|-----------------|--------------------|------|-----------------------------------------------|---|-----------------|---|----------------|---|------------------|---|------------------|---|
|                 |                    |      | Element<br>6Li                                | Q | Element<br>45Sc | Q | Element<br>89Y | Q | Element<br>103Rh | Q | Element<br>159Tb | Q |
| CCB03           | CCB03              | 1330 | 88                                            |   | 100             |   | 105            |   | 105              |   | 102              |   |

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2879

Instrument ID: P7

Start Date : 08/20/2025

Run Number: LB136893

End Date : 08/20/2025

| Lab<br>SampleID | Client<br>SampleID | Time | Internal Standard %RI For: Collision Cell |   |                |   |                  |   |                  |   |                  |   |
|-----------------|--------------------|------|-------------------------------------------|---|----------------|---|------------------|---|------------------|---|------------------|---|
|                 |                    |      | Element<br>45Sc                           | Q | Element<br>89Y | Q | Element<br>103Rh | Q | Element<br>159Tb | Q | Element<br>165Ho | Q |
| S0              | S0                 | 1049 | 100                                       |   | 100            |   | 100              |   | 100              |   | 100              |   |
| S2              | S2                 | 1056 | 101                                       |   | 99             |   | 99               |   | 100              |   | 102              |   |
| S3              | S3                 | 1059 | 98                                        |   | 97             |   | 98               |   | 98               |   | 101              |   |
| S4              | S4                 | 1102 | 100                                       |   | 99             |   | 98               |   | 99               |   | 101              |   |
| S5              | S5                 | 1105 | 99                                        |   | 97             |   | 98               |   | 99               |   | 103              |   |
| S6              | S6                 | 1108 | 100                                       |   | 98             |   | 99               |   | 101              |   | 105              |   |
| S7              | S7                 | 1111 | 104                                       |   | 103            |   | 102              |   | 105              |   | 108              |   |
| S8              | S8                 | 1114 | 105                                       |   | 105            |   | 97               |   | 103              |   | 107              |   |
| ICV01           | ICV01              | 1135 | 102                                       |   | 102            |   | 103              |   | 101              |   | 103              |   |
| LLICV01         | LLICV01            | 1150 | 98                                        |   | 99             |   | 99               |   | 99               |   | 101              |   |
| ICB01           | ICB01              | 1156 | 97                                        |   | 98             |   | 98               |   | 97               |   | 100              |   |
| ICSA01          | ICSA01             | 1200 | 100                                       |   | 100            |   | 99               |   | 100              |   | 106              |   |
| ICSAB01         | ICSAB01            | 1203 | 104                                       |   | 105            |   | 102              |   | 104              |   | 108              |   |
| CCV01           | CCV01              | 1206 | 107                                       |   | 108            |   | 103              |   | 107              |   | 111              |   |
| CCB01           | CCB01              | 1209 | 100                                       |   | 103            |   | 103              |   | 102              |   | 105              |   |
| CRI             | CRI                | 1215 | 102                                       |   | 103            |   | 103              |   | 101              |   | 106              |   |
| PB169300BL      | PB169300BL         | 1219 | 101                                       |   | 103            |   | 103              |   | 100              |   | 104              |   |
| PB169300BS      | PB169300BS         | 1222 | 103                                       |   | 104            |   | 103              |   | 104              |   | 108              |   |
| PB169300TB      | PB169300TB         | 1225 | 100                                       |   | 102            |   | 101              |   | 100              |   | 104              |   |
| Q2879-02        | OU4-TS-GRIL        | 1228 | 105                                       |   | 113            |   | 106              |   | 106              |   | 110              |   |
| Q2879-04        | OU4-TS-GRIL        | 1232 | 102                                       |   | 111            |   | 104              |   | 103              |   | 106              |   |
| Q2879-06        | OU4-TS-GRIL        | 1235 | 103                                       |   | 113            |   | 107              |   | 105              |   | 108              |   |
| Q2879-08        | OU4-TS-GRIL        | 1238 | 105                                       |   | 114            |   | 107              |   | 107              |   | 111              |   |
| Q2879-10        | OU4-TS-GRIL        | 1241 | 105                                       |   | 111            |   | 107              |   | 106              |   | 111              |   |
| Q2879-12        | OU4-TS-GRIL        | 1244 | 106                                       |   | 115            |   | 108              |   | 108              |   | 112              |   |
| CCV02           | CCV02              | 1247 | 107                                       |   | 108            |   | 104              |   | 108              |   | 112              |   |
| CCB02           | CCB02              | 1252 | 99                                        |   | 102            |   | 102              |   | 101              |   | 104              |   |
| Q2879-14        | OU4-TS-GRIL        | 1256 | 104                                       |   | 112            |   | 107              |   | 106              |   | 111              |   |
| Q2879-16        | OU4-TS-GRIL        | 1259 | 104                                       |   | 115            |   | 108              |   | 106              |   | 111              |   |
| Q2879-18        | OU4-TS-GRIL        | 1302 | 103                                       |   | 115            |   | 108              |   | 107              |   | 111              |   |
| Q2879-18DUP     | OU4-TS-GRIL        | 1305 | 102                                       |   | 114            |   | 107              |   | 107              |   | 110              |   |
| Q2879-18L       | OU4-TS-GRIL        | 1308 | 102                                       |   | 107            |   | 106              |   | 105              |   | 109              |   |
| Q2879-18MS      | OU4-TS-GRIL        | 1311 | 105                                       |   | 114            |   | 107              |   | 107              |   | 112              |   |
| Q2879-18MSD     | OU4-TS-GRIL        | 1314 | 104                                       |   | 113            |   | 107              |   | 106              |   | 110              |   |
| Q2879-18A       | OU4-TS-GRIL        | 1317 | 104                                       |   | 114            |   | 110              |   | 108              |   | 112              |   |
| CCV03           | CCV03              | 1327 | 106                                       |   | 109            |   | 104              |   | 109              |   | 114              |   |

Internal Standard %RI Limit: 30 - 120

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2879

Instrument ID: P7

Start Date : 08/20/2025

Run Number: LB136893

End Date : 08/20/2025

| Lab<br>SampleID | Client<br>SampleID | Time | Internal Standard %RI For: Collision Cell |   |         |   |         |   |         |   |         |   |
|-----------------|--------------------|------|-------------------------------------------|---|---------|---|---------|---|---------|---|---------|---|
|                 |                    |      | Element                                   |   | Element |   | Element |   | Element |   | Element |   |
|                 |                    |      | 45Sc                                      | Q | 89Y     | Q | 103Rh   | Q | 159Tb   | Q | 165Ho   | Q |
| CCB03           | CCB03              | 1330 | 99                                        |   | 103     |   | 103     |   | 103     |   | 106     |   |