

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2879

Instrument ID: P7

Start Date : 08/20/2025

Run Number: LB136893

End Date : 08/20/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
S0	S0	1049	100		100							
S2	S2	1056	100		102							
S3	S3	1059	98		100							
S4	S4	1102	101		102							
S5	S5	1105	100		100							
S6	S6	1108	104		104							
S7	S7	1111	106		106							
S8	S8	1114	104		97							
ICV01	ICV01	1135	100		100							
LLICV01	LLICV01	1150	99		98							
ICB01	ICB01	1156	99		98							
ICSA01	ICSA01	1200	103		99							
ICSAB01	ICSAB01	1203	106		102							
CCV01	CCV01	1206	111		103							
CCB01	CCB01	1209	103		101							
CRI	CRI	1215	103		101							
PB169300BL	PB169300BL	1219	101		99							
PB169300BS	PB169300BS	1222	105		102							
PB169300TB	PB169300TB	1225	102		100							
Q2879-02	OU4-TS-GRILI	1228	106		101							
Q2879-04	OU4-TS-GRILI	1232	104		99							
Q2879-06	OU4-TS-GRILI	1235	105		100							
Q2879-08	OU4-TS-GRILI	1238	105		101							
Q2879-10	OU4-TS-GRILI	1241	104		100							
Q2879-12	OU4-TS-GRILI	1244	106		101							
CCV02	CCV02	1247	108		101							
CCB02	CCB02	1252	102		99							
Q2879-14	OU4-TS-GRILI	1256	107		102							
Q2879-16	OU4-TS-GRILI	1259	109		103							
Q2879-18	OU4-TS-GRILI	1302	107		101							
Q2879-18DUP	OU4-TS-GRILI	1305	105		99							
Q2879-18L	OU4-TS-GRILI	1308	104		100							
Q2879-18MS	OU4-TS-GRILI	1311	108		102							
Q2879-18MSD	OU4-TS-GRILI	1314	107		101							

Internal Standard %RI Limit: 30 -120

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2879

Instrument ID: P7

Start Date : 08/20/2025

Run Number: LB136893

End Date : 08/20/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
Q2879-18A	OU4-TS-GRILI	1317	106		103							
CCV03	CCV03	1327	110		99							
CCB03	CCB03	1330	103		99							

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2879

Instrument ID: P7

Start Date : 08/20/2025

Run Number: LB136893

End Date : 08/20/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
S0	S0	1049	100									
S2	S2	1056	101									
S3	S3	1059	100									
S4	S4	1102	100									
S5	S5	1105	101									
S6	S6	1108	101									
S7	S7	1111	103									
S8	S8	1114	96									
ICV01	ICV01	1135	102									
LLICV01	LLICV01	1150	100									
ICB01	ICB01	1156	98									
ICSA01	ICSA01	1200	98									
ICSAB01	ICSAB01	1203	102									
CCV01	CCV01	1206	102									
CCB01	CCB01	1209	101									
CRI	CRI	1215	102									
PB169300BL	PB169300BL	1219	101									
PB169300BS	PB169300BS	1222	103									
PB169300TB	PB169300TB	1225	101									
Q2879-02	OU4-TS-GRILI	1228	104									
Q2879-04	OU4-TS-GRILI	1232	101									
Q2879-06	OU4-TS-GRILI	1235	103									
Q2879-08	OU4-TS-GRILI	1238	106									
Q2879-10	OU4-TS-GRILI	1241	106									
Q2879-12	OU4-TS-GRILI	1244	106									
CCV02	CCV02	1247	102									
CCB02	CCB02	1252	100									
Q2879-14	OU4-TS-GRILI	1256	106									
Q2879-16	OU4-TS-GRILI	1259	105									
Q2879-18	OU4-TS-GRILI	1302	104									
Q2879-18DUP	OU4-TS-GRILI	1305	104									
Q2879-18L	OU4-TS-GRILI	1308	103									
Q2879-18MS	OU4-TS-GRILI	1311	105									
Q2879-18MSD	OU4-TS-GRILI	1314	103									

Internal Standard %RI Limit: 30 -120

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2879

Instrument ID: P7

Start Date : 08/20/2025

Run Number: LB136893

End Date : 08/20/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
Q2879-18A	OU4-TS-GRILI	1317	104									
CCV03	CCV03	1327	102									
CCB03	CCB03	1330	102									