

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2879

Continuing Calib Date: 08/20/2025

Initial Calibration Date(s): 08/07/2025

08/07/2025

Continuing Calib Time: 01:10

Initial Calibration Time(s): 12:02

13:38

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.47	7.48	7.38	7.58	0.01
2,4-DCAA	7.29	7.30	7.20	7.40	0.01
Dalapon	2.67	2.67	2.57	2.77	0.00
DICHLORPROP	8.18	8.19	8.09	8.29	0.01
2,4-D	8.41	8.42	8.32	8.52	0.01
2,4,5-TP(Silvex)	9.29	9.30	9.20	9.40	0.01
2,4,5-T	9.58	9.60	9.50	9.70	0.02
2,4-DB	10.16	10.17	10.07	10.27	0.01
Dinoseb	11.36	11.37	11.27	11.47	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2879

Continuing Calib Date: 08/20/2025

Initial Calibration Date(s): 08/07/2025

08/07/2025

Continuing Calib Time: 01:10

Initial Calibration Time(s): 12:02

13:38

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.94	7.95	7.85	8.05	0.01
2,4-DCAA	7.74	7.75	7.65	7.85	0.01
Dalapon	2.70	2.70	2.60	2.80	0.01
DICHLORPROP	8.66	8.67	8.57	8.77	0.01
2,4-D	9.00	9.01	8.91	9.11	0.01
2,4,5-TP(Silvex)	9.89	9.90	9.80	10.00	0.01
2,4,5-T	10.32	10.33	10.23	10.43	0.01
2,4-DB	10.89	10.90	10.80	11.00	0.01
Dinoseb	11.27	11.28	11.18	11.38	0.01