

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** NOBI03  
**Lab Code:** ACE **SDG NO.:** Q2879  
**GC Column:** RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/07/2025 08/07/2025

**Client Sample No.:** CCAL01 **Date Analyzed:** 08/19/2025  
**Lab Sample No.:** HSTDCCC750 **Data File :** PS031494.D **Time Analyzed:** 13:09

| COMPOUND         | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|------------------|--------|-----------|--------|--------------------|-------------------|-------|
|                  |        | FROM      | TO     |                    |                   |       |
| 2,4,5-T          | 9.586  | 9.495     | 9.695  | 731.090            | 712.500           | 2.6   |
| 2,4,5-TP(Silvex) | 9.290  | 9.200     | 9.400  | 713.410            | 712.500           | 0.1   |
| 2,4-D            | 8.417  | 8.324     | 8.524  | 675.620            | 705.000           | -4.2  |
| 2,4-DB           | 10.163 | 10.070    | 10.270 | 712.340            | 712.500           | 0.0   |
| 2,4-DCAA         | 7.288  | 7.195     | 7.395  | 731.100            | 750.000           | -2.5  |
| Dalapon          | 2.671  | 2.574     | 2.774  | 593.580            | 682.500           | -13.0 |
| DICAMBA          | 7.473  | 7.381     | 7.581  | 685.160            | 705.000           | -2.8  |
| DICHLORPROP      | 8.181  | 8.089     | 8.289  | 676.850            | 705.000           | -4.0  |
| Dinoseb          | 11.361 | 11.273    | 11.473 | 709.380            | 705.000           | 0.6   |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** NOBI03  
**Lab Code:** ACE **SDG NO.:** Q2879  
**GC Column:** RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/07/2025 08/07/2025

**Client Sample No.:** CCAL01 **Date Analyzed:** 08/19/2025  
**Lab Sample No.:** HSTDCCC750 **Data File :** PS031494.D **Time Analyzed:** 13:09

| COMPOUND         | RT     | RT WINDOW<br>FROM TO |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|------------------|--------|----------------------|--------|--------------------|-------------------|-------|
| 2,4,5-T          | 10.324 | 10.231               | 10.431 | 654.170            | 712.500           | -8.2  |
| 2,4,5-TP(Silvex) | 9.895  | 9.803                | 10.003 | 668.200            | 712.500           | -6.2  |
| 2,4-D            | 9.001  | 8.906                | 9.106  | 618.620            | 705.000           | -12.3 |
| 2,4-DB           | 10.892 | 10.799               | 10.999 | 623.370            | 712.500           | -12.5 |
| 2,4-DCAA         | 7.744  | 7.649                | 7.849  | 680.760            | 750.000           | -9.2  |
| Dalapon          | 2.693  | 2.595                | 2.795  | 562.690            | 682.500           | -17.6 |
| DICAMBA          | 7.940  | 7.847                | 8.047  | 652.060            | 705.000           | -7.5  |
| DICHLORPROP      | 8.659  | 8.566                | 8.766  | 642.110            | 705.000           | -8.9  |
| Dinoseb          | 11.269 | 11.178               | 11.378 | 664.320            | 705.000           | -5.8  |