

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2879
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/07/2025 08/07/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/19/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031503.D **Time Analyzed:** 20:21

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.585	9.495	9.695	778.900	712.500	9.3
2,4,5-TP(Silvex)	9.290	9.200	9.400	751.080	712.500	5.4
2,4-D	8.414	8.324	8.524	758.260	705.000	7.6
2,4-DB	10.160	10.070	10.270	853.390	712.500	19.8
2,4-DCAA	7.287	7.195	7.395	757.340	750.000	1.0
Dalapon	2.674	2.574	2.774	620.730	682.500	-9.1
DICAMBA	7.473	7.381	7.581	710.460	705.000	0.8
DICHLORPROP	8.180	8.089	8.289	706.910	705.000	0.3
Dinoseb	11.362	11.273	11.473	749.210	705.000	6.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2879
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/07/2025 08/07/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/19/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031503.D **Time Analyzed:** 20:21

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.323	10.231	10.431	705.030	712.500	-1.0
2,4,5-TP(Silvex)	9.895	9.803	10.003	696.910	712.500	-2.2
2,4-D	8.998	8.906	9.106	627.280	705.000	-11.0
2,4-DB	10.890	10.799	10.999	660.280	712.500	-7.3
2,4-DCAA	7.742	7.649	7.849	710.850	750.000	-5.2
Dalapon	2.694	2.595	2.795	608.670	682.500	-10.8
DICAMBA	7.940	7.847	8.047	680.000	705.000	-3.5
DICHLORPROP	8.659	8.566	8.766	670.210	705.000	-4.9
Dinoseb	11.269	11.178	11.378	690.250	705.000	-2.1