

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2879
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/07/2025 08/07/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/20/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031513.D **Time Analyzed:** 01:10

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.584	9.495	9.695	774.880	712.500	8.8
2,4,5-TP(Silvex)	9.289	9.200	9.400	757.530	712.500	6.3
2,4-D	8.413	8.324	8.524	763.980	705.000	8.4
2,4-DB	10.158	10.070	10.270	852.980	712.500	19.7
2,4-DCAA	7.286	7.195	7.395	759.190	750.000	1.2
Dalapon	2.672	2.574	2.774	606.950	682.500	-11.1
DICAMBA	7.472	7.381	7.581	715.380	705.000	1.5
DICHLORPROP	8.180	8.089	8.289	713.350	705.000	1.2
Dinoseb	11.360	11.273	11.473	707.640	705.000	0.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2879
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/07/2025 08/07/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/20/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031513.D **Time Analyzed:** 01:10

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.322	10.231	10.431	723.820	712.500	1.6
2,4,5-TP(Silvex)	9.894	9.803	10.003	713.290	712.500	0.1
2,4-D	8.998	8.906	9.106	647.970	705.000	-8.1
2,4-DB	10.889	10.799	10.999	672.680	712.500	-5.6
2,4-DCAA	7.742	7.649	7.849	727.700	750.000	-3.0
Dalapon	2.695	2.595	2.795	610.970	682.500	-10.5
DICAMBA	7.940	7.847	8.047	698.840	705.000	-0.9
DICHLORPROP	8.659	8.566	8.766	690.320	705.000	-2.1
Dinoseb	11.268	11.178	11.378	653.890	705.000	-7.2