

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2879
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/07/2025 08/07/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/20/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031516.D **Time Analyzed:** 09:05

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.586	9.495	9.695	762.620	712.500	7.0
2,4,5-TP(Silvex)	9.290	9.200	9.400	737.700	712.500	3.5
2,4-D	8.416	8.324	8.524	719.390	705.000	2.0
2,4-DB	10.162	10.070	10.270	760.140	712.500	6.7
2,4-DCAA	7.288	7.195	7.395	742.540	750.000	-1.0
Dalapon	2.675	2.574	2.774	587.520	682.500	-13.9
DICAMBA	7.473	7.381	7.581	702.000	705.000	-0.4
DICHLORPROP	8.181	8.089	8.289	696.680	705.000	-1.2
Dinoseb	11.362	11.273	11.473	721.990	705.000	2.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2879
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/07/2025 08/07/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/20/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031516.D **Time Analyzed:** 09:05

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.323	10.231	10.431	692.100	712.500	-2.9
2,4,5-TP(Silvex)	9.895	9.803	10.003	696.460	712.500	-2.3
2,4-D	8.999	8.906	9.106	638.840	705.000	-9.4
2,4-DB	10.891	10.799	10.999	645.000	712.500	-9.5
2,4-DCAA	7.743	7.649	7.849	695.720	750.000	-7.2
Dalapon	2.695	2.595	2.795	574.460	682.500	-15.8
DICAMBA	7.940	7.847	8.047	675.790	705.000	-4.1
DICHLORPROP	8.659	8.566	8.766	671.040	705.000	-4.8
Dinoseb	11.269	11.178	11.378	664.260	705.000	-5.8