

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2879
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/07/2025 08/07/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/20/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031525.D **Time Analyzed:** 15:11

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.584	9.495	9.695	778.380	712.500	9.2
2,4,5-TP(Silvex)	9.289	9.200	9.400	758.820	712.500	6.5
2,4-D	8.413	8.324	8.524	759.360	705.000	7.7
2,4-DB	10.158	10.070	10.270	830.730	712.500	16.6
2,4-DCAA	7.287	7.195	7.395	758.060	750.000	1.1
Dalapon	2.676	2.574	2.774	603.560	682.500	-11.6
DICAMBA	7.472	7.381	7.581	714.140	705.000	1.3
DICHLORPROP	8.179	8.089	8.289	713.880	705.000	1.3
Dinoseb	11.361	11.273	11.473	735.980	705.000	4.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2879
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/07/2025 08/07/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/20/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031525.D **Time Analyzed:** 15:11

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.322	10.231	10.431	721.580	712.500	1.3
2,4,5-TP(Silvex)	9.894	9.803	10.003	709.930	712.500	-0.4
2,4-D	8.997	8.906	9.106	628.810	705.000	-10.8
2,4-DB	10.889	10.799	10.999	670.320	712.500	-5.9
2,4-DCAA	7.742	7.649	7.849	726.610	750.000	-3.1
Dalapon	2.694	2.595	2.795	588.910	682.500	-13.7
DICAMBA	7.940	7.847	8.047	699.360	705.000	-0.8
DICHLORPROP	8.658	8.566	8.766	667.460	705.000	-5.3
Dinoseb	11.268	11.178	11.378	676.610	705.000	-4.0