

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS080725

Review By	Abdul	Review On	8/8/2025 8:24:12 AM
Supervise By	mohammad	Supervise On	8/11/2025 1:55:16 AM
SubDirectory	PS080725	HP Acquire Method	HP Processing Method ps080725 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24562
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS031377.D	07 Aug 2025 11:14		AR\AJ	Ok
2	I.BLK	I.BLK	PS031378.D	07 Aug 2025 11:38		AR\AJ	Ok
3	HSTDICC200	HSTDICC200	PS031379.D	07 Aug 2025 12:02		AR\AJ	Ok
4	HSTDICC500	HSTDICC500	PS031380.D	07 Aug 2025 12:26		AR\AJ	Ok
5	HSTDICC750	HSTDICC750	PS031381.D	07 Aug 2025 12:50		AR\AJ	Ok
6	HSTDICC1000	HSTDICC1000	PS031382.D	07 Aug 2025 13:14		AR\AJ	Ok
7	HSTDICC1500	HSTDICC1500	PS031383.D	07 Aug 2025 13:38		AR\AJ	Ok,M
8	HSTDICV750	ICVPS080725	PS031384.D	07 Aug 2025 14:02		AR\AJ	Ok
9	I.BLK	I.BLK	PS031385.D	07 Aug 2025 14:27	Looks like hexan	AR\AJ	Not Ok
10	HSTDCCC750	HSTDCCC750	PS031386.D	07 Aug 2025 14:51		AR\AJ	Ok
11	PB169160BL	PB169160BL	PS031387.D	07 Aug 2025 16:03	Opening IBLK fail	AR\AJ	Not Ok
12	PB169160BS	PB169160BS	PS031388.D	07 Aug 2025 16:27	Opening IBLK fail	AR\AJ	Not Ok
13	PB169160BSD	PB169160BSD	PS031389.D	07 Aug 2025 16:51	Opening IBLK fail	AR\AJ	Not Ok
14	Q2759-04	LIQUID-DRUM	PS031390.D	07 Aug 2025 17:15	Opening IBLK fail	AR\AJ	Not Ok
15	I.BLK	I.BLK	PS031391.D	07 Aug 2025 17:40	Opening IBLK fail	AR\AJ	Not Ok
16	HSTDCCC750	HSTDCCC750	PS031392.D	07 Aug 2025 18:04		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS081925

Review By	Abdul	Review On	8/20/2025 8:43:36 AM
Supervise By	mohammad	Supervise On	8/21/2025 1:32:01 AM
SubDirectory	PS081925	HP Acquire Method	HP Processing Method ps080725 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24802,PP24556,PP24557,PP24558,PP24559,PP24560,PP24803,PP24804
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24804
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS031492.D	19 Aug 2025 12:21		AR\AJ	Ok
2	I.BLK	I.BLK	PS031493.D	19 Aug 2025 12:45		AR\AJ	Ok
3	HSTDCCC750	HSTDCCC750	PS031494.D	19 Aug 2025 13:09		AR\AJ	Ok
4	PB169287BL	PB169287BL	PS031495.D	19 Aug 2025 17:08		AR\AJ	Ok
5	PB169287BS	PB169287BS	PS031496.D	19 Aug 2025 17:32		AR\AJ	Ok,M
6	Q2879-01	OU4-TS-GRILLO-TSCF	PS031497.D	19 Aug 2025 17:56		AR\AJ	Ok
7	Q2879-03	OU4-TS-GRILLO-TSCF	PS031498.D	19 Aug 2025 18:20	Surrogate low in 1st column	AR\AJ	ReRun
8	Q2879-05	OU4-TS-GRILLO-TSCF	PS031499.D	19 Aug 2025 18:44		AR\AJ	Ok
9	Q2879-07	OU4-TS-GRILLO-TSCF	PS031500.D	19 Aug 2025 19:08		AR\AJ	Ok
10	Q2879-09	OU4-TS-GRILLO-TSCF	PS031501.D	19 Aug 2025 19:33	Surrogate low in both column	AR\AJ	Not Ok
11	I.BLK	I.BLK	PS031502.D	19 Aug 2025 19:57		AR\AJ	Ok
12	HSTDCCC750	HSTDCCC750	PS031503.D	19 Aug 2025 20:21		AR\AJ	Ok,M
13	Q2879-11	OU4-TS-GRILLO-TSCF	PS031504.D	19 Aug 2025 21:33	Surrogate low in both column	AR\AJ	Not Ok
14	Q2879-13	OU4-TS-GRILLO-TSCF	PS031505.D	19 Aug 2025 21:57	Surrogate low in 1st column	AR\AJ	ReRun
15	Q2879-15	OU4-TS-GRILLO-TSCF	PS031506.D	19 Aug 2025 22:22		AR\AJ	Ok,M
16	Q2879-17	OU4-TS-GRILLO-TSCF	PS031507.D	19 Aug 2025 22:46	Surrogate low in both column	AR\AJ	Not Ok
17	Q2884-03	302	PS031508.D	19 Aug 2025 23:10		AR\AJ	Ok,M
18	Q2888-01	VNJ-210	PS031509.D	19 Aug 2025 23:34		AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS081925

Review By	Abdul	Review On	8/20/2025 8:43:36 AM
Supervise By	mohammad	Supervise On	8/21/2025 1:32:01 AM
SubDirectory	PS081925	HP Acquire Method	HP Processing Method ps080725 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24802,PP24556,PP24557,PP24558,PP24559,PP24560,PP24803,PP24804		
CCC	PP24559		
Internal Standard/PEM	PP24804		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q2888-01MS	VNJ-210MS	PS031510.D	19 Aug 2025 23:58	some compound recovery fail	AR\AJ	Not Ok
20	Q2888-01MSD	VNJ-210MSD	PS031511.D	20 Aug 2025 00:22	F Flag in comp#14 , some compound recovery fail , RPD fail	AR\AJ	Not Ok
21	I.BLK	I.BLK	PS031512.D	20 Aug 2025 00:46		AR\AJ	Ok
22	HSTDCCC750	HSTDCCC750	PS031513.D	20 Aug 2025 01:10		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS082025

Review By	Abdul	Review On	8/21/2025 9:37:30 AM
Supervise By		Supervise On	
SubDirectory	PS082025	HP Acquire Method	HP Processing Method ps080725 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24802,PP24556,PP24557,PP24558,PP24559,PP24560,PP24803,PP24804
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24804
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS031514.D	20 Aug 2025 08:17		AR\AJ	Ok
2	I.BLK	I.BLK	PS031515.D	20 Aug 2025 08:41		AR\AJ	Ok
3	HSTDCCC750	HSTDCCC750	PS031516.D	20 Aug 2025 09:05		AR\AJ	Ok,NS
4	Q2879-03RE	OU4-TS-GRILLO-TSCF	PS031517.D	20 Aug 2025 09:29	Surrogate low in 1st column	AR\AJ	Confirms
5	Q2879-09	OU4-TS-GRILLO-TSCF	PS031518.D	20 Aug 2025 09:53		AR\AJ	Ok
6	Q2879-11	OU4-TS-GRILLO-TSCF	PS031519.D	20 Aug 2025 10:21		AR\AJ	Ok,NS
7	Q2879-13RE	OU4-TS-GRILLO-TSCF	PS031520.D	20 Aug 2025 10:45	Surrogate low in 1st column	AR\AJ	Confirms
8	Q2879-17	OU4-TS-GRILLO-TSCF	PS031521.D	20 Aug 2025 11:10		AR\AJ	Ok,NS
9	Q2888-01MS	VNJ-210MS	PS031522.D	20 Aug 2025 11:59		AR\AJ	Ok,NS
10	Q2888-01MSD	VNJ-210MSD	PS031523.D	20 Aug 2025 12:24		AR\AJ	Ok,NS
11	I.BLK	I.BLK	PS031524.D	20 Aug 2025 14:47		AR\AJ	Ok
12	HSTDCCC750	HSTDCCC750	PS031525.D	20 Aug 2025 15:11		AR\AJ	Ok,NS

M : Manual Integration