

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2879

Analytical Method: 8151A

Client: Nobis Group

Datafile : PS031496.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB169287BS (Column 1)	DICAMBA	166.5	184	ug/Kg	111				38	132	
	Dalapon	166.5	156	ug/Kg	94				70	130	
	DICHLORPROP	166.5	185	ug/Kg	111				28	155	
	2,4-D	166.5	186	ug/Kg	112				28	144	
	2,4,5-TP(Silvex)	166.5	191	ug/Kg	115				43	129	
	2,4,5-T	166.5	191	ug/Kg	115				31	138	
	2,4-DB	166.5	178	ug/Kg	107				34	142	
	Dinoseb	166.5	190	ug/Kg	114				57	152	
	DICAMBA	166.5	172	ug/Kg	103				38	132	
PB169287BS (Column 2)	Dalapon	166.5	150	ug/Kg	90				70	130	
	DICHLORPROP	166.5	172	ug/Kg	103				28	155	
	2,4-D	166.5	175	ug/Kg	105				28	144	
	2,4,5-TP(Silvex)	166.5	178	ug/Kg	107				43	129	
	2,4,5-T	166.5	171	ug/Kg	103				31	138	
	2,4-DB	166.5	156	ug/Kg	94				34	142	
	Dinoseb	166.5	176	ug/Kg	106				57	152	