

Instrument ID: BNA\_F

**Daily Analysis Runlog For Sequence/QC Batch ID # BF081525**

|              |          |                      |                      |
|--------------|----------|----------------------|----------------------|
| Review By    | Rahul    | Review On            | 8/18/2025 8:39:29 AM |
| Supervise By | Jagrut   | Supervise On         | 8/18/2025 8:43:38 AM |
| SubDirectory | BF081525 | HP Acquire Method    | BNA_F                |
|              |          | HP Processing Method | bf081525             |

| STD. NAME                | STD REF.#                                               |
|--------------------------|---------------------------------------------------------|
| Tune/Reschk              | SP6757                                                  |
| Initial Calibration Stds | SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864 |
| CCC                      | SP6860                                                  |
| Internal Standard/PEM    | S13170,10ul/1000ul sample                               |
| ICV/I.BLK                | SP6770                                                  |
| Surrogate Standard       |                                                         |
| MS/MSD Standard          |                                                         |
| LCS Standard             |                                                         |

| Sr# | SampleID    | ClientID     | Data File Name | Date-Time         | Comment                                 | Operator | Status |
|-----|-------------|--------------|----------------|-------------------|-----------------------------------------|----------|--------|
| 1   | DFTPP       | DFTPP        | BF143415.D     | 15 Aug 2025 13:59 |                                         | RC/JU    | Ok     |
| 2   | SSTDICC2.5  | SSTDICC2.5   | BF143416.D     | 15 Aug 2025 14:29 |                                         | RC/JU    | Ok     |
| 3   | SSTDICC005  | SSTDICC005   | BF143417.D     | 15 Aug 2025 14:58 |                                         | RC/JU    | Ok     |
| 4   | SSTDICC010  | SSTDICC010   | BF143418.D     | 15 Aug 2025 15:28 |                                         | RC/JU    | Ok,M   |
| 5   | SSTDICC020  | SSTDICC020   | BF143419.D     | 15 Aug 2025 15:58 | Compound#41,54 Kept on LR               | RC/JU    | Ok,M   |
| 6   | SSTDICCC040 | SSTDICCC040  | BF143420.D     | 15 Aug 2025 16:28 | Benzaldehyde failed in the Calibration. | RC/JU    | Ok     |
| 7   | SSTDICC050  | SSTDICC050   | BF143421.D     | 15 Aug 2025 16:58 |                                         | RC/JU    | Ok     |
| 8   | SSTDICC060  | SSTDICC060   | BF143422.D     | 15 Aug 2025 17:27 |                                         | RC/JU    | Ok     |
| 9   | SSTDICC080  | SSTDICC080   | BF143423.D     | 15 Aug 2025 17:57 | compound #77 removed.                   | RC/JU    | Ok     |
| 10  | SSTDICV040  | ICVBF081525  | BF143424.D     | 15 Aug 2025 18:57 |                                         | RC/JU    | Ok     |
| 11  | PB169229BL  | PB169229BL   | BF143425.D     | 15 Aug 2025 19:27 |                                         | RC/JU    | Ok     |
| 12  | PB169229BS  | PB169229BS   | BF143426.D     | 15 Aug 2025 19:56 |                                         | RC/JU    | Ok,M   |
| 13  | SSTDCCC040  | SSTDCCC040EC | BF143427.D     | 15 Aug 2025 20:26 |                                         | RC/JU    | Ok     |

M : Manual Integration

Instrument ID: BNA\_F

**Daily Analysis Runlog For Sequence/QC Batch ID # BF081925**

|              |          |                      |                      |
|--------------|----------|----------------------|----------------------|
| Review By    | Rahul    | Review On            | 8/20/2025 8:37:42 AM |
| Supervise By | Jagrut   | Supervise On         | 8/20/2025 1:25:01 PM |
| SubDirectory | BF081925 | HP Acquire Method    | BNA_F                |
|              |          | HP Processing Method | bf081525             |

| STD. NAME                                                                                          | STD REF.#                                                         |
|----------------------------------------------------------------------------------------------------|-------------------------------------------------------------------|
| Tune/Reschk<br>Initial Calibration Stds                                                            | SP6757<br>SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864 |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | SP6860<br>S13171,10ul/1000ul sample<br>SP6770                     |

| Sr# | SampleID    | ClientID            | Data File Name | Date-Time         | Comment                                      | Operator | Status |
|-----|-------------|---------------------|----------------|-------------------|----------------------------------------------|----------|--------|
| 1   | DFTPP       | DFTPP               | BF143444.D     | 19 Aug 2025 08:39 |                                              | RC/JU    | Ok     |
| 2   | SSTDCCC040  | SSTDCCC040          | BF143445.D     | 19 Aug 2025 09:36 |                                              | RC/JU    | Ok     |
| 3   | PB169285BL  | PB169285BL          | BF143446.D     | 19 Aug 2025 10:05 | Used only for the DOD projects in the batch. | RC/JU    | Ok     |
| 4   | PB169285BS  | PB169285BS          | BF143447.D     | 19 Aug 2025 10:34 | Used only for the DOD projects in the batch. | RC/JU    | Ok,M   |
| 5   | Q2873-01    | DRILL CUTTING 1-3 C | BF143448.D     | 19 Aug 2025 11:06 | Method failed for com. #9                    | RC/JU    | Not Ok |
| 6   | Q2873-01MS  | DRILL CUTTING 1-3 C | BF143449.D     | 19 Aug 2025 11:35 | Used only for the DOD projects in the batch. | RC/JU    | Ok     |
| 7   | Q2873-01MSD | DRILL CUTTING 1-3 C | BF143450.D     | 19 Aug 2025 12:04 | Used only for the DOD projects in the batch. | RC/JU    | Ok     |
| 8   | Q2879-01    | OU4-TS-GRILLO-TSCF  | BF143451.D     | 19 Aug 2025 12:33 |                                              | RC/JU    | Ok,M   |
| 9   | Q2879-03    | OU4-TS-GRILLO-TSCF  | BF143452.D     | 19 Aug 2025 13:01 |                                              | RC/JU    | Ok,M   |
| 10  | Q2879-05    | OU4-TS-GRILLO-TSCF  | BF143453.D     | 19 Aug 2025 13:30 |                                              | RC/JU    | Ok,M   |
| 11  | Q2879-07    | OU4-TS-GRILLO-TSCF  | BF143454.D     | 19 Aug 2025 13:59 |                                              | RC/JU    | Ok,M   |
| 12  | Q2879-09    | OU4-TS-GRILLO-TSCF  | BF143455.D     | 19 Aug 2025 14:28 |                                              | RC/JU    | Ok,M   |
| 13  | Q2879-11    | OU4-TS-GRILLO-TSCF  | BF143456.D     | 19 Aug 2025 14:57 |                                              | RC/JU    | Ok,M   |
| 14  | Q2879-13    | OU4-TS-GRILLO-TSCF  | BF143457.D     | 19 Aug 2025 15:26 | Surrogates failed.                           | RC/JU    | Ok,M   |
| 15  | Q2879-15    | OU4-TS-GRILLO-TSCF  | BF143458.D     | 19 Aug 2025 15:55 |                                              | RC/JU    | Ok,M   |
| 16  | Q2879-17    | OU4-TS-GRILLO-TSCF  | BF143459.D     | 19 Aug 2025 16:24 |                                              | RC/JU    | Ok,M   |
| 17  | Q2888-02    | VNJ-210             | BF143460.D     | 19 Aug 2025 16:58 |                                              | RC/JU    | Ok     |

Instrument ID: BNA\_F

**Daily Analysis Runlog For Sequence/QC Batch ID # BF081925**

|                          |          |                                                         |                      |                      |          |
|--------------------------|----------|---------------------------------------------------------|----------------------|----------------------|----------|
| Review By                | Rahul    | Review On                                               | 8/20/2025 8:37:42 AM |                      |          |
| Supervise By             | Jagrut   | Supervise On                                            | 8/20/2025 1:25:01 PM |                      |          |
| SubDirectory             | BF081925 | HP Acquire Method                                       | BNA_F                | HP Processing Method | bf081525 |
| STD. NAME                |          | STD REF.#                                               |                      |                      |          |
| Tune/Reschk              |          | SP6757                                                  |                      |                      |          |
| Initial Calibration Stds |          | SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864 |                      |                      |          |
| CCC                      |          | SP6860                                                  |                      |                      |          |
| Internal Standard/PEM    |          | S13171,10ul/1000ul sample                               |                      |                      |          |
| ICV/I.BLK                |          | SP6770                                                  |                      |                      |          |
| Surrogate Standard       |          |                                                         |                      |                      |          |
| MS/MSD Standard          |          |                                                         |                      |                      |          |
| LCS Standard             |          |                                                         |                      |                      |          |

|    |             |              |            |                   |  |       |      |
|----|-------------|--------------|------------|-------------------|--|-------|------|
| 18 | Q2891-01    | BARRICADE    | BF143461.D | 19 Aug 2025 17:27 |  | RC/JU | Ok   |
| 19 | Q2891-02    | WALL-1       | BF143462.D | 19 Aug 2025 17:57 |  | RC/JU | Ok   |
| 20 | Q2891-03    | WALL-2       | BF143463.D | 19 Aug 2025 18:26 |  | RC/JU | Ok   |
| 21 | Q2891-04    | WALL-3       | BF143464.D | 19 Aug 2025 18:55 |  | RC/JU | Ok   |
| 22 | SSTDCCC040  | SSTDCCC040EC | BF143465.D | 19 Aug 2025 19:53 |  | RC/JU | Ok   |
| 23 | DFTPP       | DFTPP        | BF143466.D | 19 Aug 2025 20:52 |  | RC/JU | Ok   |
| 24 | SSTDCCC040  | SSTDCCC040   | BF143467.D | 19 Aug 2025 21:21 |  | RC/JU | Ok   |
| 25 | PB169312BL  | PB169312BL   | BF143468.D | 19 Aug 2025 21:50 |  | RC/JU | Ok   |
| 26 | PB169312BS  | PB169312BS   | BF143469.D | 19 Aug 2025 22:19 |  | RC/JU | Ok,M |
| 27 | PB169291TB  | PB169291TB   | BF143470.D | 19 Aug 2025 22:48 |  | RC/JU | Ok   |
| 28 | Q2888-02MS  | VNJ-210MS    | BF143471.D | 19 Aug 2025 23:17 |  | RC/JU | Ok,M |
| 29 | Q2888-02MSD | VNJ-210MSD   | BF143472.D | 19 Aug 2025 23:46 |  | RC/JU | Ok,M |
| 30 | Q2683-07    | DSN003       | BF143473.D | 20 Aug 2025 00:15 |  | RC/JU | Ok   |
| 31 | Q2683-03    | DSN001       | BF143474.D | 20 Aug 2025 00:45 |  | RC/JU | Ok   |
| 32 | Q2884-04    | 302          | BF143475.D | 20 Aug 2025 01:14 |  | RC/JU | Ok   |

M : Manual Integration