

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2879

Client ID : SSTDCCC040

Date Analyzed: 08/19/2025

Lab File ID: BF143445.D

Time Analyzed: 09:36

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	123862	6.928	470823	8.21	251926	9.96
UPPER LIMIT	247724	7.428	941646	8.71	503852	10.463
LOWER LIMIT	61931	6.428	235412	7.71	125963	9.463
EPA SAMPLE NO.						
01 PB169285BL	133029	6.93	514311	8.20	286595	9.96
02 PB169285BS	132148	6.93	508774	8.21	275188	9.96
03 DRILL CUTTING 1-3 COMFMS	121157	6.93	450837	8.21	228619	9.96
04 DRILL CUTTING 1-3 COMFMSD	124558	6.93	459157	8.21	257230	9.96
05 OU4-TS-GRILLO-TSCP03-081425	112068	6.92	363238	8.20	160000	9.96
06 OU4-TS-GRILLO-TSCP04-081425	111871	6.93	347487	8.20	150668	9.96
07 OU4-TS-GRILLO-TSCP05-081425	93170	6.93	315157	8.20	143358	9.96
08 OU4-TS-GRILLO-TSCP06-081425	93502	6.93	282085	8.20	137092	9.96
09 OU4-TS-GRILLO-TSCP07-081425	81535	6.93	266043	8.20	139719	9.96
10 OU4-TS-GRILLO-TSCP08-081425	98829	6.93	307720	8.21	157233	9.96
11 OU4-TS-GRILLO-TSCP09-081425	111617	6.93	365896	8.21	170198	9.96
12 OU4-TS-GRILLO-TSCP10-081425	101637	6.93	308767	8.21	162569	9.96
13 OU4-TS-GRILLO-TSCP11-081425	97604	6.93	324465	8.21	169368	9.96

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.