



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: NOBI03
Lab Code: ACE SDG No.: Q2879
Instrument ID: MSVOA_Y Calibration Date/Time: 08/19/2025 19:08
Lab File ID: VY023160.D Init. Calib. Date(s): 08/12/2025 08/12/2025
Heated Purge: (Y/N) Y Init. Calib. Time(s): 14:54 16:47
GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Methylene Chloride	50.000	50.25	-0.4940033	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.