

Surrogate Summary

SDG No.: **Q2879**

Client: **Nobis Group**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2879-01	OU4-TS-GRILLO-TSCP03-081425	1,2-Dichloroethane-d4	50	56.2	112		71	136
		Dibromofluoromethane	50	53.5	107		78	119
		Toluene-d8	50	50.4	101		85	116
		4-Bromofluorobenzene	50	39.6	79		79	119
Q2879-03	OU4-TS-GRILLO-TSCP04-081425	1,2-Dichloroethane-d4	50	53.6	107		71	136
		Dibromofluoromethane	50	52.3	105		78	119
		Toluene-d8	50	50.3	101		85	116
		4-Bromofluorobenzene	50	42.4	85		79	119
Q2879-05	OU4-TS-GRILLO-TSCP05-081425	1,2-Dichloroethane-d4	50	54.3	109		71	136
		Dibromofluoromethane	50	54.3	109		78	119
		Toluene-d8	50	50.1	100		85	116
		4-Bromofluorobenzene	50	42.1	84		79	119
Q2879-07	OU4-TS-GRILLO-TSCP06-081425	1,2-Dichloroethane-d4	50	54.6	109		71	136
		Dibromofluoromethane	50	52.0	104		78	119
		Toluene-d8	50	50.2	100		85	116
		4-Bromofluorobenzene	50	39.6	79		79	119
Q2879-09	OU4-TS-GRILLO-TSCP07-081425	1,2-Dichloroethane-d4	50	56.1	112		71	136
		Dibromofluoromethane	50	53.7	107		78	119
		Toluene-d8	50	50.6	101		85	116
		4-Bromofluorobenzene	50	41.8	84		79	119
Q2879-11	OU4-TS-GRILLO-TSCP08-081425	1,2-Dichloroethane-d4	50	55.5	111		71	136
		Dibromofluoromethane	50	53.0	106		78	119
		Toluene-d8	50	51.1	102		85	116
		4-Bromofluorobenzene	50	42.8	86		79	119
Q2879-13	OU4-TS-GRILLO-TSCP09-081425	1,2-Dichloroethane-d4	50	60.2	120		71	136
		Dibromofluoromethane	50	55.2	110		78	119
		Toluene-d8	50	51.5	103		85	116
		4-Bromofluorobenzene	50	43.8	88		79	119
Q2879-15	OU4-TS-GRILLO-TSCP10-081425	1,2-Dichloroethane-d4	50	56.0	112		71	136
		Dibromofluoromethane	50	53.6	107		78	119
		Toluene-d8	50	51.0	102		85	116
		4-Bromofluorobenzene	50	42.0	84		79	119
Q2879-17	OU4-TS-GRILLO-TSCP11-081425	1,2-Dichloroethane-d4	50	54.9	110		71	136
		Dibromofluoromethane	50	52.6	105		78	119
		Toluene-d8	50	50.4	101		85	116
		4-Bromofluorobenzene	50	38.1	76	*	79	119
Q2879-17RE	OU4-TS-GRILLO-TSCP11-081425RE	1,2-Dichloroethane-d4	50	60.2	120		71	136
		Dibromofluoromethane	50	59.0	118		78	119
		Toluene-d8	50	49.8	100		85	116
		4-Bromofluorobenzene	50	38.2	76	*	79	119
VY0819SBL01	VY0819SBL01	1,2-Dichloroethane-d4	50	53.4	107		71	136
		Dibromofluoromethane	50	52.4	105		78	119
		Toluene-d8	50	51.3	103		85	116
		4-Bromofluorobenzene	50	49.5	99		79	119
VY0819SBS01	VY0819SBS01	1,2-Dichloroethane-d4	50	47.4	95		71	136
		Dibromofluoromethane	50	48.6	97		78	119
		Toluene-d8	50	49.6	99		85	116
		4-Bromofluorobenzene	50	48.5	97		79	119
VY0819SBSD01	VY0819SBSD01	1,2-Dichloroethane-d4	50	46.1	92		71	136
		Dibromofluoromethane	50	47.6	95		78	119

Surrogate Summary

SDG No.: Q2879

Client: Nobis Group

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
VY0819SBSD01	VY0819SBSD01	Toluene-d8	50	48.5	97		85	116
		4-Bromofluorobenzene	50	46.8	94		79	119
VY0820SBL01	VY0820SBL01	1,2-Dichloroethane-d4	50	56.3	112		71	136
		Dibromofluoromethane	50	53.1	106		78	119
		Toluene-d8	50	51.9	104		85	116
		4-Bromofluorobenzene	50	46.8	94		79	119
VY0820SBS01	VY0820SBS01	1,2-Dichloroethane-d4	50	47.5	95		71	136
		Dibromofluoromethane	50	49.7	99		78	119
		Toluene-d8	50	49.7	99		85	116
		4-Bromofluorobenzene	50	47.6	95		79	119