

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>NOBI03</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2879</u>
Lab File ID:	<u>VY023162.D</u>	Date Analyzed:	<u>08/20/2025</u>
Instrument ID:	<u>MSVOA_Y</u>	Time Analyzed:	<u>09:26</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>Y</u>

	IS4 AREA #	RT #				
12 HOUR STD	347499	13.347				
UPPER LIMIT	694998	13.847				
LOWER LIMIT	173750	12.847				
EPA SAMPLE NO.						
OU4-TS-GRILLO-TSCP04-081425	234966	13.35				
OU4-TS-GRILLO-TSCP05-081425	241273	13.35				
OU4-TS-GRILLO-TSCP07-081425	248201	13.35				
OU4-TS-GRILLO-TSCP08-081425	250948	13.35				
OU4-TS-GRILLO-TSCP09-081425	295759	13.35				
OU4-TS-GRILLO-TSCP10-081425	235287	13.35				
OU4-TS-GRILLO-TSCP11-081425RE	77921 *	13.35				
VY0820SBL01	382880	13.35				
VY0820SBS01	323804	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.