

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2879

Lab File ID: VY023136.D

BFB Injection Date: 08/19/2025

Instrument ID: MSVOA_Y

BFB Injection Time: 07:45

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.1
75	30.0 - 60.0% of mass 95	51.4
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	0.8 (1) 1
174	50.0 - 100.0% of mass 95	87.6
175	5.0 - 9.0% of mass 174	6.6 (7.5) 1
176	95.0 - 101.0% of mass 174	84.1 (96.1) 1
177	5.0 - 9.0% of mass 176	5.4 (6.4) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY023137.D	08/19/2025	09:11
VY0819SBL01	VY0819SBL01	VY023138.D	08/19/2025	09:49
VY0819SBS01	VY0819SBS01	VY023139.D	08/19/2025	10:19
VY0819SBSD01	VY0819SBSD01	VY023140.D	08/19/2025	10:42
OU4-TS-GRILLO-TSCP03-081425	Q2879-01	VY023148.D	08/19/2025	14:04
OU4-TS-GRILLO-TSCP06-081425	Q2879-07	VY023151.D	08/19/2025	15:15
OU4-TS-GRILLO-TSCP11-081425	Q2879-17	VY023156.D	08/19/2025	17:12
VSTDCCC050EC	VSTDCCC050	VY023160.D	08/19/2025	19:08