

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: CHEM02
 Lab Code: ACE SDG No.: Q2880
 Instrument ID: MSVOA_X Calibration Date/Time: 08/15/2025 10:22
 Lab File ID: VX047351.D Init. Calib. Date(s): 08/15/2025 08/15/2025
 Heated Purge: (Y/N) N Init. Calib. Time(s): 09:40 12:30
 GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
2-Chloroethyl Vinyl ether	250.000	269.74	-7.895203	20