

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CHEM02
 Lab Code: ACE SDG NO.: Q2880
 Lab File ID: VX047351.D Date Analyzed: 08/15/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:22
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	238847	5.57	405572	6.77	363128	10.06
UPPER LIMIT	477694	6.068	811144	7.269	726256	10.555
LOWER LIMIT	119424	5.068	202786	6.269	181564	9.555
EPA SAMPLE NO.						
STORAGE-BLANK-SOIL-REF-6	298008	5.57	567332	6.77	570443	10.06
STORAGE-BLANK-WATER-REF-5	319933	5.56	613476	6.77	606257	10.06
STORAGE-BLANK-WATER-REF-4	304017	5.57	589868	6.77	577258	10.06
STORAGE-BLANK-SAMPLE-MANAGEMENT	316568	5.57	608089	6.78	598972	10.06

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.