

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN037617.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q2881
DFTPP Injection Date: 08/20/2025
DFTPP Injection Time: 03:35

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.4 (0.9) 1
69	Mass 69 relative abundance	39.2
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
365	Greater than 1% of mass 198	3.6
441	Present, but less than mass 443	11.8
442	Greater than 50% of mass 198	78.3
443	15.0 - 24.0% of mass 442	14.9 (19) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037618.D	08/20/2025	04:14
PB169286BL	PB169286BL	BN037619.D	08/20/2025	04:50
RW8-SP100-20250814	Q2881-01	BN037623.D	08/20/2025	07:13
RW8-SP303-20250814	Q2881-02	BN037624.D	08/20/2025	07:49
PB169286BS	PB169286BS	BN037630.D	08/20/2025	11:25
SSTDCCC0.4EC	SSTDCCC0.4	BN037631.D	08/20/2025	14:56