

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2882

Client ID : SSTDCCC0.4

Date Analyzed: 08/20/2025

Lab File ID: BN037618.D

Time Analyzed: 04:14

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2622	7.71	6166	10.50	3094	14.35
UPPER LIMIT	5244	8.21	12332	10.998	6188	14.845
LOWER LIMIT	1311	7.21	3083	9.998	1547	13.845
EPA SAMPLE NO.						
01 RW7-SP100-20250814	2353	7.71	5433	10.50	2588	14.35
02 RW7-SP201-20250814	2063	7.71	4828	10.50	2403	14.35
03 RW7-SP302-20250814	2158	7.71	5059	10.50	2475	14.35
04 RW7-SP303-20250814	2442	7.71	5851	10.49	3334	14.35
05 PB169286BL	2323	7.71	5229	10.50	2344	14.35
06 PB169286BS	2389	7.71	5616	10.49	2629	14.35

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.