

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2885

Client ID : SSTDCCC0.4

Date Analyzed: 08/20/2025

Lab File ID: BN037633.D

Time Analyzed: 16:57

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2754	7.71	6418	10.49	3102	14.35
UPPER LIMIT	5508	8.21	12836	10.987	6204	14.845
LOWER LIMIT	1377	7.21	3209	9.987	1551	13.845
EPA SAMPLE NO.						
01 PB169328BL	2138	7.71	4908	10.49	2089	14.35
02 PB169328BS	2188	7.71	5030	10.49	2230	14.35
03 PB169328BSD	2346	7.71	5382	10.49	2393	14.35
04 RW5-SP100-20250814	2105	7.71	4906	10.49	2332	14.35
05 RW5-SP201-20250814	1991	7.71	4567	10.49	2148	14.35
06 RW5-SP303-20250814	1820	7.71	4111	10.49	1992	14.35

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.