

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2885

Client ID : SSTDCCC0.4

Date Analyzed: 08/21/2025

Lab File ID: BN037643.D

Time Analyzed: 09:51

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2301	7.71	5449	10.49	2612	14.35
UPPER LIMIT	4602	8.21	10898	10.987	5224	14.845
LOWER LIMIT	1150.5	7.21	2724.5	9.987	1306	13.845
EPA SAMPLE NO.						
01 RW5-SP100-20250814DL	3001	7.71	6896	10.49	3211	14.35

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.