

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2886</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Tetra Tech NUS, Inc.</u>	Datafile :	<u>VX047360.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0815WBS01	Dichlorodifluoromethane	20	17.0	ug/L	85			32	152	
	Chloromethane	20	16.0	ug/L	80			50	139	
	Vinyl chloride	20	16.4	ug/L	82			58	137	
	Bromomethane	20	18.0	ug/L	90			53	141	
	Chloroethane	20	17.2	ug/L	86			60	138	
	Trichlorofluoromethane	20	16.5	ug/L	83			65	141	
	1,1,2-Trichlorotrifluoroethane	20	17.1	ug/L	86			70	136	
	1,1-Dichloroethene	20	16.8	ug/L	84			71	131	
	Acetone	100	97.4	ug/L	97			39	160	
	Carbon disulfide	20	16.1	ug/L	81			64	133	
	Methyl tert-butyl Ether	20	17.1	ug/L	86			71	124	
	Methyl Acetate	20	17.7	ug/L	89			56	136	
	Methylene Chloride	20	16.7	ug/L	84			74	124	
	trans-1,2-Dichloroethene	20	16.8	ug/L	84			75	124	
	1,1-Dichloroethane	20	16.7	ug/L	84			77	125	
	Cyclohexane	20	16.4	ug/L	82			71	130	
	2-Butanone	100	93.4	ug/L	93			56	143	
	Carbon Tetrachloride	20	16.8	ug/L	84			72	136	
	cis-1,2-Dichloroethene	20	16.4	ug/L	82			78	123	
	Bromochloromethane	20	20.3	ug/L	102			78	123	
	Chloroform	20	17.2	ug/L	86			79	124	
	1,1,1-Trichloroethane	20	16.8	ug/L	84			74	131	
	Methylcyclohexane	20	16.9	ug/L	85			72	132	
	Benzene	20	17.3	ug/L	86			79	120	
	1,2-Dichloroethane	20	17.6	ug/L	88			73	128	
	Trichloroethene	20	17.1	ug/L	86			79	123	
	1,2-Dichloropropane	20	17.2	ug/L	86			78	122	
	Bromodichloromethane	20	17.2	ug/L	86			79	125	
	4-Methyl-2-Pentanone	100	92.0	ug/L	92			67	130	
	Toluene	20	17.6	ug/L	88			80	121	
	t-1,3-Dichloropropene	20	16.9	ug/L	85			73	127	
	cis-1,3-Dichloropropene	20	17.2	ug/L	86			75	124	
	1,1,2-Trichloroethane	20	17.7	ug/L	89			80	119	
	2-Hexanone	100	92.3	ug/L	92			57	139	
	Dibromochloromethane	20	17.0	ug/L	85			74	126	
	1,2-Dibromoethane	20	17.7	ug/L	89			77	121	
	Tetrachloroethene	20	16.9	ug/L	85			74	129	
	Chlorobenzene	20	17.0	ug/L	85			82	118	
	Ethyl Benzene	20	17.1	ug/L	86			79	121	
	m/p-Xylenes	40	34.7	ug/L	87			80	121	
	o-Xylene	20	17.1	ug/L	86			78	122	
	Styrene	20	17.8	ug/L	89			78	123	
	Bromoform	20	16.9	ug/L	85			66	130	
	Isopropylbenzene	20	17.0	ug/L	85			72	131	
	1,1,2,2-Tetrachloroethane	20	17.5	ug/L	88			71	121	
	1,3-Dichlorobenzene	20	17.0	ug/L	85			80	119	
	1,4-Dichlorobenzene	20	16.8	ug/L	84			79	118	
	1,2-Dichlorobenzene	20	17.0	ug/L	85			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2886</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Tetra Tech NUS, Inc.</u>	Datafile :	<u>VX047360.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0815WBS01	1,2-Dibromo-3-Chloropropane	20	16.9	ug/L	85			62	128	
	1,2,4-Trichlorobenzene	20	16.5	ug/L	83			69	130	
	1,2,3-Trichlorobenzene	20	16.7	ug/L	84			69	129	