

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2886</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Tetra Tech NUS, Inc.</u>	Datafile :	<u>VX047361.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0815WBSD01	Dichlorodifluoromethane	20	17.1	ug/L	86	1		32	152	20
	Chloromethane	20	16.2	ug/L	81	1		50	139	20
	Vinyl chloride	20	16.5	ug/L	83	1		58	137	20
	Bromomethane	20	19.2	ug/L	96	6		53	141	20
	Chloroethane	20	16.3	ug/L	81	6		60	138	20
	Trichlorofluoromethane	20	16.1	ug/L	81	2		65	141	20
	1,1,2-Trichlorotrifluoroethane	20	17.1	ug/L	86	0		70	136	20
	1,1-Dichloroethene	20	16.3	ug/L	81	4		71	131	20
	Acetone	100	91.2	ug/L	91	6		39	160	20
	Carbon disulfide	20	16.2	ug/L	81	0		64	133	20
	Methyl tert-butyl Ether	20	18.0	ug/L	90	5		71	124	20
	Methyl Acetate	20	18.0	ug/L	90	1		56	136	20
	Methylene Chloride	20	17.6	ug/L	88	5		74	124	20
	trans-1,2-Dichloroethene	20	16.7	ug/L	84	0		75	124	20
	1,1-Dichloroethane	20	17.3	ug/L	86	2		77	125	20
	Cyclohexane	20	16.6	ug/L	83	1		71	130	20
	2-Butanone	100	92.2	ug/L	92	1		56	143	20
	Carbon Tetrachloride	20	16.5	ug/L	83	1		72	136	20
	cis-1,2-Dichloroethene	20	16.9	ug/L	85	4		78	123	20
	Bromochloromethane	20	21.4	ug/L	107	5		78	123	20
	Chloroform	20	17.8	ug/L	89	3		79	124	20
	1,1,1-Trichloroethane	20	17.2	ug/L	86	2		74	131	20
	Methylcyclohexane	20	16.7	ug/L	84	1		72	132	20
	Benzene	20	17.3	ug/L	86	0		79	120	20
	1,2-Dichloroethane	20	17.8	ug/L	89	1		73	128	20
	Trichloroethene	20	16.6	ug/L	83	4		79	123	20
	1,2-Dichloropropane	20	17.0	ug/L	85	1		78	122	20
	Bromodichloromethane	20	17.6	ug/L	88	2		79	125	20
	4-Methyl-2-Pentanone	100	93.8	ug/L	94	2		67	130	20
	Toluene	20	17.6	ug/L	88	0		80	121	20
	t-1,3-Dichloropropene	20	17.8	ug/L	89	5		73	127	20
	cis-1,3-Dichloropropene	20	17.7	ug/L	89	3		75	124	20
	1,1,2-Trichloroethane	20	18.2	ug/L	91	2		80	119	20
	2-Hexanone	100	94.0	ug/L	94	2		57	139	20
	Dibromochloromethane	20	17.4	ug/L	87	2		74	126	20
	1,2-Dibromoethane	20	17.8	ug/L	89	0		77	121	20
	Tetrachloroethene	20	16.9	ug/L	85	0		74	129	20
	Chlorobenzene	20	17.1	ug/L	86	1		82	118	20
	Ethyl Benzene	20	17.0	ug/L	85	1		79	121	20
	m/p-Xylenes	40	34.5	ug/L	86	1		80	121	20
	o-Xylene	20	17.2	ug/L	86	0		78	122	20
	Styrene	20	17.6	ug/L	88	1		78	123	20
	Bromoform	20	17.6	ug/L	88	3		66	130	20
	Isopropylbenzene	20	16.8	ug/L	84	1		72	131	20
	1,1,2,2-Tetrachloroethane	20	18.1	ug/L	91	3		71	121	20
	1,3-Dichlorobenzene	20	17.1	ug/L	86	1		80	119	20
	1,4-Dichlorobenzene	20	17.1	ug/L	86	2		79	118	20
	1,2-Dichlorobenzene	20	17.4	ug/L	87	2		80	119	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2886</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Tetra Tech NUS, Inc.</u>	Datafile :	<u>VX047361.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0815WBSD01	1,2-Dibromo-3-Chloropropane	20	17.5	ug/L	88	3		62	128	20
	1,2,4-Trichlorobenzene	20	17.2	ug/L	86	4		69	130	20
	1,2,3-Trichlorobenzene	20	17.2	ug/L	86	2		69	129	20