

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG NO.: Q2886
 Lab File ID: VX047351.D Date Analyzed: 08/15/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:22
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	238847	5.57	405572	6.77	363128	10.06
UPPER LIMIT	477694	6.068	811144	7.269	726256	10.555
LOWER LIMIT	119424	5.068	202786	6.269	181564	9.555
EPA SAMPLE NO.						
BP-TT182D-TB-20250814	322553	5.57	632578	6.77	621493	10.06
VX0815WBL01	213035	5.56	413649	6.76	396719	10.06
VX0815WBS01	251230	5.56	424536	6.77	385995	10.06
VX0815WBSD01	232894	5.56	402024	6.76	366622	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.