

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TETR06
Lab Code: ACE SDG NO.: Q2886
Lab File ID: VX047371.D Date Analyzed: 08/18/2025
Instrument ID: MSVOA_X Time Analyzed: 10:00
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	269622	5.56	444434	6.76	389564	10.06
UPPER LIMIT	539244	6.056	888868	7.263	779128	10.555
LOWER LIMIT	134811	5.056	222217	6.263	194782	9.555
EPA SAMPLE NO.						
BP-TT182D-GW-20250814	318653	5.57	625670	6.77	611655	10.06
VX0818WBL01	325331	5.56	630884	6.76	620229	10.06
VX0818WBS01	256302	5.56	426348	6.77	389002	10.06

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.