

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG NO.: Q2886
 Lab File ID: VX047351.D Date Analyzed: 08/15/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:22
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	177366	12.018				
UPPER LIMIT	354732	12.518				
LOWER LIMIT	88683	11.518				
EPA SAMPLE NO.						
BP-TT182D-TB-20250814	310393	12.02				
VX0815WBL01	194794	12.02				
VX0815WBS01	195233	12.02				
VX0815WBSD01	184649	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.