

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q2886

Lab File ID: VX047347.D

BFB Injection Date: 08/15/2025

Instrument ID: MSVOA_X

BFB Injection Time: 08:28

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.3
75	30.0 - 60.0% of mass 95	51.4
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	0.9 (1.1) 1
174	50.0 - 100.0% of mass 95	74.3
175	5.0 - 9.0% of mass 174	6.1 (8.2) 1
176	95.0 - 101.0% of mass 174	71.9 (96.8) 1
177	5.0 - 9.0% of mass 176	4.4 (6.2) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VX047349.D	08/15/2025	09:40
VSTDIC020	VSTDIC020	VX047350.D	08/15/2025	10:01
VSTDIC050	VSTDIC050	VX047351.D	08/15/2025	10:22
VSTDIC100	VSTDIC100	VX047352.D	08/15/2025	10:43
VSTDIC150	VSTDIC150	VX047353.D	08/15/2025	11:05
VSTDIC001	VSTDIC001	VX047356.D	08/15/2025	12:30
VX0815WBL01	VX0815WBL01	VX047358.D	08/15/2025	13:39
VX0815WBS01	VX0815WBS01	VX047360.D	08/15/2025	14:21
VX0815WBSD01	VX0815WBSD01	VX047361.D	08/15/2025	14:49
BP-TT182D-TB-20250814	Q2886-01	VX047366.D	08/15/2025	16:36
VSTDCCC050EC	VSTDCCC050	VX047369.D	08/15/2025	17:41