

Matrix Spike Summary

Client:	BAPS North Bergen Development Inc.	SDG No.:	Q2892
Project:	BAPS 2000 Tonnelle Ave	Sample ID:	Q2879-17
Client ID:	OU4-TS-GRILLO-TSCP11-081425MS	Percent Solids for Spike Sample:	77

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/Kg	75-125	2.20		0.15	J	2.5	1	82		08/19/2025

Matrix Spike Summary

Client:	BAPS North Bergen Development Inc.	SDG No.:	Q2892
Project:	BAPS 2000 Tonnelle Ave	Sample ID:	Q2879-17
Client ID:	OU4-TS-GRILLO-TSCP11-081425MSD	Percent Solids for Spike Sample:	77

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/Kg	75-125	2.40		0.15	J	2.6	1	87		08/19/2025