

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: BAPS North Bergen Development Inc.

SDG No.: Q2892

Contract: BAPS03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV22	Mercury	3.86	4.0	96	90 - 110	CV	08/19/2025	11:44	LB136873

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: BAPS North Bergen Development Inc.

SDG No.: Q2892

Contract: BAPS03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV68	Mercury	4.93	5.0	99	90 - 110	CV	08/19/2025	11:49	LB136873
CCV69	Mercury	4.89	5.0	98	90 - 110	CV	08/19/2025	12:19	LB136873
CCV70	Mercury	4.68	5.0	94	90 - 110	CV	08/19/2025	12:55	LB136873

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: BAPS North Bergen Development Inc.

SDG No.: Q2892

Contract: BAPS03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7850	8000	98	90 - 110	P	08/19/2025	12:17	LB136886
	Antimony	4070	4000	102	90 - 110	P	08/19/2025	12:17	LB136886
	Arsenic	3990	4000	100	90 - 110	P	08/19/2025	12:17	LB136886
	Barium	7940	8000	99	90 - 110	P	08/19/2025	12:17	LB136886
	Beryllium	209	200	104	90 - 110	P	08/19/2025	12:17	LB136886
	Cadmium	1970	2000	98	90 - 110	P	08/19/2025	12:17	LB136886
	Calcium	19600	20000	98	90 - 110	P	08/19/2025	12:17	LB136886
	Chromium	803	800	100	90 - 110	P	08/19/2025	12:17	LB136886
	Cobalt	1980	2000	99	90 - 110	P	08/19/2025	12:17	LB136886
	Copper	1020	1000	102	90 - 110	P	08/19/2025	12:17	LB136886
	Iron	4010	4000	100	90 - 110	P	08/19/2025	12:17	LB136886
	Lead	3990	4000	100	90 - 110	P	08/19/2025	12:17	LB136886
	Magnesium	19600	20000	98	90 - 110	P	08/19/2025	12:17	LB136886
	Manganese	1980	2000	99	90 - 110	P	08/19/2025	12:17	LB136886
	Nickel	1980	2000	99	90 - 110	P	08/19/2025	12:17	LB136886
	Potassium	19600	20000	98	90 - 110	P	08/19/2025	12:17	LB136886
	Selenium	4030	4000	101	90 - 110	P	08/19/2025	12:17	LB136886
	Silver	1010	1000	101	90 - 110	P	08/19/2025	12:17	LB136886
	Sodium	19600	20000	98	90 - 110	P	08/19/2025	12:17	LB136886
	Thallium	3920	4000	98	90 - 110	P	08/19/2025	12:17	LB136886
	Vanadium	1970	2000	98	90 - 110	P	08/19/2025	12:17	LB136886
	Zinc	2000	2000	100	90 - 110	P	08/19/2025	12:17	LB136886

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: BAPS North Bergen Development Inc.

SDG No.: Q2892

Contract: BAPS03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	117	100	117	80 - 120	P	08/19/2025	12:23	LB136886
	Antimony	55.2	50.0	110	80 - 120	P	08/19/2025	12:23	LB136886
	Arsenic	20.1	20.0	100	80 - 120	P	08/19/2025	12:23	LB136886
	Barium	100	100	100	80 - 120	P	08/19/2025	12:23	LB136886
	Beryllium	6.28	6.0	105	80 - 120	P	08/19/2025	12:23	LB136886
	Cadmium	6.00	6.0	100	80 - 120	P	08/19/2025	12:23	LB136886
	Calcium	2120	2000	106	80 - 120	P	08/19/2025	12:23	LB136886
	Chromium	10.0	10.0	100	80 - 120	P	08/19/2025	12:23	LB136886
	Cobalt	30.3	30.0	101	80 - 120	P	08/19/2025	12:23	LB136886
	Copper	22.1	20.0	111	80 - 120	P	08/19/2025	12:23	LB136886
	Iron	112	100	112	80 - 120	P	08/19/2025	12:23	LB136886
	Lead	12.5	12.0	104	80 - 120	P	08/19/2025	12:23	LB136886
	Magnesium	2190	2000	109	80 - 120	P	08/19/2025	12:23	LB136886
	Manganese	22.1	20.0	111	80 - 120	P	08/19/2025	12:23	LB136886
	Nickel	40.8	40.0	102	80 - 120	P	08/19/2025	12:23	LB136886
	Potassium	1900	2000	95	80 - 120	P	08/19/2025	12:23	LB136886
	Selenium	18.3	20.0	91	80 - 120	P	08/19/2025	12:23	LB136886
	Silver	10.6	10.0	106	80 - 120	P	08/19/2025	12:23	LB136886
	Sodium	1960	2000	98	80 - 120	P	08/19/2025	12:23	LB136886
	Thallium	43.9	40.0	110	80 - 120	P	08/19/2025	12:23	LB136886
	Vanadium	43.0	40.0	108	80 - 120	P	08/19/2025	12:23	LB136886
	Zinc	44.7	40.0	112	80 - 120	P	08/19/2025	12:23	LB136886

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: BAPS North Bergen Development Inc.

SDG No.: Q2892

Contract: BAPS03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9980	10000	100	90 - 110	P	08/19/2025	12:59	LB136886
	Antimony	5080	5000	102	90 - 110	P	08/19/2025	12:59	LB136886
	Arsenic	5050	5000	101	90 - 110	P	08/19/2025	12:59	LB136886
	Barium	10100	10000	100	90 - 110	P	08/19/2025	12:59	LB136886
	Beryllium	244	250	98	90 - 110	P	08/19/2025	12:59	LB136886
	Cadmium	2480	2500	99	90 - 110	P	08/19/2025	12:59	LB136886
	Calcium	25100	25000	100	90 - 110	P	08/19/2025	12:59	LB136886
	Chromium	1000	1000	100	90 - 110	P	08/19/2025	12:59	LB136886
	Cobalt	2480	2500	99	90 - 110	P	08/19/2025	12:59	LB136886
	Copper	1260	1250	101	90 - 110	P	08/19/2025	12:59	LB136886
	Iron	5140	5000	103	90 - 110	P	08/19/2025	12:59	LB136886
	Lead	4950	5000	99	90 - 110	P	08/19/2025	12:59	LB136886
	Magnesium	24700	25000	99	90 - 110	P	08/19/2025	12:59	LB136886
	Manganese	2500	2500	100	90 - 110	P	08/19/2025	12:59	LB136886
	Nickel	2480	2500	99	90 - 110	P	08/19/2025	12:59	LB136886
	Potassium	25400	25000	102	90 - 110	P	08/19/2025	12:59	LB136886
	Selenium	5090	5000	102	90 - 110	P	08/19/2025	12:59	LB136886
	Silver	1260	1250	100	90 - 110	P	08/19/2025	12:59	LB136886
	Sodium	26000	25000	104	90 - 110	P	08/19/2025	12:59	LB136886
	Thallium	5150	5000	103	90 - 110	P	08/19/2025	12:59	LB136886
	Vanadium	2500	2500	100	90 - 110	P	08/19/2025	12:59	LB136886
	Zinc	2540	2500	102	90 - 110	P	08/19/2025	12:59	LB136886
CCV02	Aluminum	9760	10000	98	90 - 110	P	08/19/2025	14:05	LB136886
	Antimony	5010	5000	100	90 - 110	P	08/19/2025	14:05	LB136886
	Arsenic	4980	5000	100	90 - 110	P	08/19/2025	14:05	LB136886
	Barium	9800	10000	98	90 - 110	P	08/19/2025	14:05	LB136886
	Beryllium	239	250	96	90 - 110	P	08/19/2025	14:05	LB136886
	Cadmium	2430	2500	97	90 - 110	P	08/19/2025	14:05	LB136886
	Calcium	24400	25000	98	90 - 110	P	08/19/2025	14:05	LB136886
	Chromium	982	1000	98	90 - 110	P	08/19/2025	14:05	LB136886
	Cobalt	2440	2500	97	90 - 110	P	08/19/2025	14:05	LB136886
	Copper	1240	1250	99	90 - 110	P	08/19/2025	14:05	LB136886
	Iron	5080	5000	102	90 - 110	P	08/19/2025	14:05	LB136886
	Lead	4850	5000	97	90 - 110	P	08/19/2025	14:05	LB136886

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: BAPS North Bergen Development Inc.

SDG No.: Q2892

Contract: BAPS03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	24200	25000	97	90 - 110	P	08/19/2025	14:05	LB136886
	Manganese	2440	2500	98	90 - 110	P	08/19/2025	14:05	LB136886
	Nickel	2440	2500	98	90 - 110	P	08/19/2025	14:05	LB136886
	Potassium	24800	25000	99	90 - 110	P	08/19/2025	14:05	LB136886
	Selenium	5020	5000	100	90 - 110	P	08/19/2025	14:05	LB136886
	Silver	1240	1250	99	90 - 110	P	08/19/2025	14:05	LB136886
	Sodium	25000	25000	100	90 - 110	P	08/19/2025	14:05	LB136886
	Thallium	5060	5000	101	90 - 110	P	08/19/2025	14:05	LB136886
	Vanadium	2480	2500	99	90 - 110	P	08/19/2025	14:05	LB136886
	Zinc	2500	2500	100	90 - 110	P	08/19/2025	14:05	LB136886
CCV03	Aluminum	9870	10000	99	90 - 110	P	08/19/2025	14:54	LB136886
	Antimony	5060	5000	101	90 - 110	P	08/19/2025	14:54	LB136886
	Arsenic	5040	5000	101	90 - 110	P	08/19/2025	14:54	LB136886
	Barium	10000	10000	100	90 - 110	P	08/19/2025	14:54	LB136886
	Beryllium	243	250	97	90 - 110	P	08/19/2025	14:54	LB136886
	Cadmium	2470	2500	99	90 - 110	P	08/19/2025	14:54	LB136886
	Calcium	24600	25000	98	90 - 110	P	08/19/2025	14:54	LB136886
	Chromium	1000	1000	100	90 - 110	P	08/19/2025	14:54	LB136886
	Cobalt	2470	2500	99	90 - 110	P	08/19/2025	14:54	LB136886
	Copper	1250	1250	100	90 - 110	P	08/19/2025	14:54	LB136886
	Iron	5160	5000	103	90 - 110	P	08/19/2025	14:54	LB136886
	Lead	4910	5000	98	90 - 110	P	08/19/2025	14:54	LB136886
	Magnesium	24600	25000	98	90 - 110	P	08/19/2025	14:54	LB136886
	Manganese	2470	2500	99	90 - 110	P	08/19/2025	14:54	LB136886
	Nickel	2470	2500	99	90 - 110	P	08/19/2025	14:54	LB136886
	Potassium	25100	25000	100	90 - 110	P	08/19/2025	14:54	LB136886
	Selenium	5080	5000	102	90 - 110	P	08/19/2025	14:54	LB136886
	Silver	1260	1250	101	90 - 110	P	08/19/2025	14:54	LB136886
	Sodium	24900	25000	100	90 - 110	P	08/19/2025	14:54	LB136886
	Thallium	5060	5000	101	90 - 110	P	08/19/2025	14:54	LB136886
	Vanadium	2500	2500	100	90 - 110	P	08/19/2025	14:54	LB136886
	Zinc	2540	2500	102	90 - 110	P	08/19/2025	14:54	LB136886
CCV04	Aluminum	9740	10000	97	90 - 110	P	08/19/2025	15:47	LB136886
	Antimony	5020	5000	100	90 - 110	P	08/19/2025	15:47	LB136886

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: BAPS North Bergen Development Inc.

SDG No.: Q2892

Contract: BAPS03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	5000	5000	100	90 - 110	P	08/19/2025	15:47	LB136886
	Barium	9800	10000	98	90 - 110	P	08/19/2025	15:47	LB136886
	Beryllium	239	250	96	90 - 110	P	08/19/2025	15:47	LB136886
	Cadmium	2440	2500	98	90 - 110	P	08/19/2025	15:47	LB136886
	Calcium	24300	25000	97	90 - 110	P	08/19/2025	15:47	LB136886
	Chromium	990	1000	99	90 - 110	P	08/19/2025	15:47	LB136886
	Cobalt	2440	2500	98	90 - 110	P	08/19/2025	15:47	LB136886
	Copper	1240	1250	99	90 - 110	P	08/19/2025	15:47	LB136886
	Iron	5100	5000	102	90 - 110	P	08/19/2025	15:47	LB136886
	Lead	4860	5000	97	90 - 110	P	08/19/2025	15:47	LB136886
	Magnesium	24200	25000	97	90 - 110	P	08/19/2025	15:47	LB136886
	Manganese	2430	2500	97	90 - 110	P	08/19/2025	15:47	LB136886
	Nickel	2440	2500	98	90 - 110	P	08/19/2025	15:47	LB136886
	Potassium	24700	25000	99	90 - 110	P	08/19/2025	15:47	LB136886
	Selenium	5030	5000	101	90 - 110	P	08/19/2025	15:47	LB136886
	Silver	1250	1250	100	90 - 110	P	08/19/2025	15:47	LB136886
	Sodium	24400	25000	98	90 - 110	P	08/19/2025	15:47	LB136886
	Thallium	4860	5000	97	90 - 110	P	08/19/2025	15:47	LB136886
	Vanadium	2480	2500	99	90 - 110	P	08/19/2025	15:47	LB136886
	Zinc	2520	2500	101	90 - 110	P	08/19/2025	15:47	LB136886
CCV05	Aluminum	9760	10000	98	90 - 110	P	08/19/2025	16:38	LB136886
	Antimony	5070	5000	101	90 - 110	P	08/19/2025	16:38	LB136886
	Arsenic	5090	5000	102	90 - 110	P	08/19/2025	16:38	LB136886
	Barium	9760	10000	98	90 - 110	P	08/19/2025	16:38	LB136886
	Beryllium	239	250	96	90 - 110	P	08/19/2025	16:38	LB136886
	Cadmium	2470	2500	99	90 - 110	P	08/19/2025	16:38	LB136886
	Calcium	24400	25000	98	90 - 110	P	08/19/2025	16:38	LB136886
	Chromium	995	1000	100	90 - 110	P	08/19/2025	16:38	LB136886
	Cobalt	2470	2500	99	90 - 110	P	08/19/2025	16:38	LB136886
	Copper	1250	1250	100	90 - 110	P	08/19/2025	16:38	LB136886
	Iron	5100	5000	102	90 - 110	P	08/19/2025	16:38	LB136886
	Lead	4910	5000	98	90 - 110	P	08/19/2025	16:38	LB136886
	Magnesium	24400	25000	98	90 - 110	P	08/19/2025	16:38	LB136886
	Manganese	2430	2500	97	90 - 110	P	08/19/2025	16:38	LB136886

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: BAPS North Bergen Development Inc.

SDG No.: Q2892

Contract: BAPS03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2470	2500	99	90 - 110	P	08/19/2025	16:38	LB136886
	Potassium	25000	25000	100	90 - 110	P	08/19/2025	16:38	LB136886
	Selenium	5140	5000	103	90 - 110	P	08/19/2025	16:38	LB136886
	Silver	1260	1250	100	90 - 110	P	08/19/2025	16:38	LB136886
	Sodium	24900	25000	100	90 - 110	P	08/19/2025	16:38	LB136886
	Thallium	5290	5000	106	90 - 110	P	08/19/2025	16:38	LB136886
	Vanadium	2470	2500	99	90 - 110	P	08/19/2025	16:38	LB136886
	Zinc	2540	2500	102	90 - 110	P	08/19/2025	16:38	LB136886
CCV06	Aluminum	9810	10000	98	90 - 110	P	08/19/2025	17:29	LB136886
	Antimony	5120	5000	102	90 - 110	P	08/19/2025	17:29	LB136886
	Arsenic	5070	5000	102	90 - 110	P	08/19/2025	17:29	LB136886
	Barium	9990	10000	100	90 - 110	P	08/19/2025	17:29	LB136886
	Beryllium	235	250	94	90 - 110	P	08/19/2025	17:29	LB136886
	Cadmium	2460	2500	98	90 - 110	P	08/19/2025	17:29	LB136886
	Calcium	24400	25000	98	90 - 110	P	08/19/2025	17:29	LB136886
	Chromium	994	1000	99	90 - 110	P	08/19/2025	17:29	LB136886
	Cobalt	2470	2500	99	90 - 110	P	08/19/2025	17:29	LB136886
	Copper	1250	1250	100	90 - 110	P	08/19/2025	17:29	LB136886
	Iron	5290	5000	106	90 - 110	P	08/19/2025	17:29	LB136886
	Lead	4880	5000	98	90 - 110	P	08/19/2025	17:29	LB136886
	Magnesium	24300	25000	97	90 - 110	P	08/19/2025	17:29	LB136886
	Manganese	2460	2500	98	90 - 110	P	08/19/2025	17:29	LB136886
	Nickel	2460	2500	98	90 - 110	P	08/19/2025	17:29	LB136886
	Potassium	25400	25000	102	90 - 110	P	08/19/2025	17:29	LB136886
	Selenium	5150	5000	103	90 - 110	P	08/19/2025	17:29	LB136886
	Silver	1260	1250	101	90 - 110	P	08/19/2025	17:29	LB136886
	Sodium	25300	25000	101	90 - 110	P	08/19/2025	17:29	LB136886
	Thallium	5200	5000	104	90 - 110	P	08/19/2025	17:29	LB136886
	Vanadium	2490	2500	100	90 - 110	P	08/19/2025	17:29	LB136886
	Zinc	2540	2500	102	90 - 110	P	08/19/2025	17:29	LB136886
CCV07	Aluminum	9740	10000	97	90 - 110	P	08/19/2025	18:02	LB136886
	Antimony	5130	5000	103	90 - 110	P	08/19/2025	18:02	LB136886
	Arsenic	5090	5000	102	90 - 110	P	08/19/2025	18:02	LB136886
	Barium	9730	10000	97	90 - 110	P	08/19/2025	18:02	LB136886

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: BAPS North Bergen Development Inc.

SDG No.: Q2892

Contract: BAPS03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	233	250	93	90 - 110	P	08/19/2025	18:02	LB136886
	Cadmium	2450	2500	98	90 - 110	P	08/19/2025	18:02	LB136886
	Calcium	24200	25000	97	90 - 110	P	08/19/2025	18:02	LB136886
	Chromium	990	1000	99	90 - 110	P	08/19/2025	18:02	LB136886
	Cobalt	2460	2500	98	90 - 110	P	08/19/2025	18:02	LB136886
	Copper	1250	1250	100	90 - 110	P	08/19/2025	18:02	LB136886
	Iron	5200	5000	104	90 - 110	P	08/19/2025	18:02	LB136886
	Lead	4850	5000	97	90 - 110	P	08/19/2025	18:02	LB136886
	Magnesium	23900	25000	96	90 - 110	P	08/19/2025	18:02	LB136886
	Manganese	2430	2500	97	90 - 110	P	08/19/2025	18:02	LB136886
	Nickel	2450	2500	98	90 - 110	P	08/19/2025	18:02	LB136886
	Potassium	25200	25000	101	90 - 110	P	08/19/2025	18:02	LB136886
	Selenium	5150	5000	103	90 - 110	P	08/19/2025	18:02	LB136886
	Silver	1250	1250	100	90 - 110	P	08/19/2025	18:02	LB136886
	Sodium	24900	25000	100	90 - 110	P	08/19/2025	18:02	LB136886
	Thallium	5290	5000	106	90 - 110	P	08/19/2025	18:02	LB136886
	Vanadium	2470	2500	99	90 - 110	P	08/19/2025	18:02	LB136886
	Zinc	2530	2500	101	90 - 110	P	08/19/2025	18:02	LB136886

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: BAPS North Bergen Development Inc.

SDG No.: Q2892

Contract: BAPS03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7800	8000	98	90 - 110	P	08/20/2025	13:04	LB136898
	Antimony	4060	4000	102	90 - 110	P	08/20/2025	13:04	LB136898
	Arsenic	3950	4000	99	90 - 110	P	08/20/2025	13:04	LB136898
	Barium	7740	8000	97	90 - 110	P	08/20/2025	13:04	LB136898
	Beryllium	217	200	109	90 - 110	P	08/20/2025	13:04	LB136898
	Cadmium	1960	2000	98	90 - 110	P	08/20/2025	13:04	LB136898
	Calcium	19400	20000	97	90 - 110	P	08/20/2025	13:04	LB136898
	Chromium	809	800	101	90 - 110	P	08/20/2025	13:04	LB136898
	Cobalt	1970	2000	99	90 - 110	P	08/20/2025	13:04	LB136898
	Copper	1010	1000	101	90 - 110	P	08/20/2025	13:04	LB136898
	Iron	3970	4000	99	90 - 110	P	08/20/2025	13:04	LB136898
	Lead	3970	4000	99	90 - 110	P	08/20/2025	13:04	LB136898
	Magnesium	19400	20000	97	90 - 110	P	08/20/2025	13:04	LB136898
	Manganese	1950	2000	97	90 - 110	P	08/20/2025	13:04	LB136898
	Nickel	1970	2000	98	90 - 110	P	08/20/2025	13:04	LB136898
	Potassium	19600	20000	98	90 - 110	P	08/20/2025	13:04	LB136898
	Selenium	3990	4000	100	90 - 110	P	08/20/2025	13:04	LB136898
	Silver	1010	1000	101	90 - 110	P	08/20/2025	13:04	LB136898
	Sodium	18900	20000	94	90 - 110	P	08/20/2025	13:04	LB136898
	Thallium	3920	4000	98	90 - 110	P	08/20/2025	13:04	LB136898
	Vanadium	1940	2000	97	90 - 110	P	08/20/2025	13:04	LB136898
	Zinc	2010	2000	101	90 - 110	P	08/20/2025	13:04	LB136898

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: BAPS North Bergen Development Inc.

SDG No.: Q2892

Contract: BAPS03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	105	100	105	80 - 120	P	08/20/2025	13:17	LB136898
	Antimony	49.3	50.0	98	80 - 120	P	08/20/2025	13:17	LB136898
	Arsenic	20.0	20.0	100	80 - 120	P	08/20/2025	13:17	LB136898
	Barium	97.7	100	98	80 - 120	P	08/20/2025	13:17	LB136898
	Beryllium	6.15	6.0	103	80 - 120	P	08/20/2025	13:17	LB136898
	Cadmium	5.60	6.0	93	80 - 120	P	08/20/2025	13:17	LB136898
	Calcium	2050	2000	102	80 - 120	P	08/20/2025	13:17	LB136898
	Chromium	9.70	10.0	97	80 - 120	P	08/20/2025	13:17	LB136898
	Cobalt	28.7	30.0	96	80 - 120	P	08/20/2025	13:17	LB136898
	Copper	20.8	20.0	104	80 - 120	P	08/20/2025	13:17	LB136898
	Iron	103	100	103	80 - 120	P	08/20/2025	13:17	LB136898
	Lead	12.0	12.0	100	80 - 120	P	08/20/2025	13:17	LB136898
	Magnesium	2160	2000	108	80 - 120	P	08/20/2025	13:17	LB136898
	Manganese	21.9	20.0	109	80 - 120	P	08/20/2025	13:17	LB136898
	Nickel	39.1	40.0	98	80 - 120	P	08/20/2025	13:17	LB136898
	Potassium	1870	2000	94	80 - 120	P	08/20/2025	13:17	LB136898
	Selenium	21.2	20.0	106	80 - 120	P	08/20/2025	13:17	LB136898
	Silver	10.4	10.0	104	80 - 120	P	08/20/2025	13:17	LB136898
	Sodium	1910	2000	96	80 - 120	P	08/20/2025	13:17	LB136898
	Thallium	41.1	40.0	103	80 - 120	P	08/20/2025	13:17	LB136898
	Vanadium	42.6	40.0	107	80 - 120	P	08/20/2025	13:17	LB136898
	Zinc	41.2	40.0	103	80 - 120	P	08/20/2025	13:17	LB136898

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: BAPS North Bergen Development Inc.

SDG No.: Q2892

Contract: BAPS03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9920	10000	99	90 - 110	P	08/20/2025	13:47	LB136898
	Antimony	4980	5000	100	90 - 110	P	08/20/2025	13:47	LB136898
	Arsenic	4980	5000	100	90 - 110	P	08/20/2025	13:47	LB136898
	Barium	9810	10000	98	90 - 110	P	08/20/2025	13:47	LB136898
	Beryllium	246	250	98	90 - 110	P	08/20/2025	13:47	LB136898
	Cadmium	2470	2500	99	90 - 110	P	08/20/2025	13:47	LB136898
	Calcium	24800	25000	99	90 - 110	P	08/20/2025	13:47	LB136898
	Chromium	1010	1000	101	90 - 110	P	08/20/2025	13:47	LB136898
	Cobalt	2460	2500	98	90 - 110	P	08/20/2025	13:47	LB136898
	Copper	1250	1250	100	90 - 110	P	08/20/2025	13:47	LB136898
	Iron	5020	5000	100	90 - 110	P	08/20/2025	13:47	LB136898
	Lead	4900	5000	98	90 - 110	P	08/20/2025	13:47	LB136898
	Magnesium	24700	25000	99	90 - 110	P	08/20/2025	13:47	LB136898
	Manganese	2470	2500	99	90 - 110	P	08/20/2025	13:47	LB136898
	Nickel	2460	2500	98	90 - 110	P	08/20/2025	13:47	LB136898
	Potassium	24800	25000	99	90 - 110	P	08/20/2025	13:47	LB136898
	Selenium	5010	5000	100	90 - 110	P	08/20/2025	13:47	LB136898
	Silver	1240	1250	100	90 - 110	P	08/20/2025	13:47	LB136898
	Sodium	25000	25000	100	90 - 110	P	08/20/2025	13:47	LB136898
	Thallium	4970	5000	100	90 - 110	P	08/20/2025	13:47	LB136898
	Vanadium	2480	2500	99	90 - 110	P	08/20/2025	13:47	LB136898
	Zinc	2520	2500	101	90 - 110	P	08/20/2025	13:47	LB136898
CCV02	Aluminum	9850	10000	98	90 - 110	P	08/20/2025	14:39	LB136898
	Antimony	4990	5000	100	90 - 110	P	08/20/2025	14:39	LB136898
	Arsenic	4960	5000	99	90 - 110	P	08/20/2025	14:39	LB136898
	Barium	9760	10000	98	90 - 110	P	08/20/2025	14:39	LB136898
	Beryllium	243	250	97	90 - 110	P	08/20/2025	14:39	LB136898
	Cadmium	2450	2500	98	90 - 110	P	08/20/2025	14:39	LB136898
	Calcium	24500	25000	98	90 - 110	P	08/20/2025	14:39	LB136898
	Chromium	1000	1000	100	90 - 110	P	08/20/2025	14:39	LB136898
	Cobalt	2440	2500	98	90 - 110	P	08/20/2025	14:39	LB136898
	Copper	1240	1250	100	90 - 110	P	08/20/2025	14:39	LB136898
	Iron	5060	5000	101	90 - 110	P	08/20/2025	14:39	LB136898
	Lead	4860	5000	97	90 - 110	P	08/20/2025	14:39	LB136898

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: BAPS North Bergen Development Inc.

SDG No.: Q2892

Contract: BAPS03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	24200	25000	97	90 - 110	P	08/20/2025	14:39	LB136898
	Manganese	2440	2500	98	90 - 110	P	08/20/2025	14:39	LB136898
	Nickel	2450	2500	98	90 - 110	P	08/20/2025	14:39	LB136898
	Potassium	25300	25000	101	90 - 110	P	08/20/2025	14:39	LB136898
	Selenium	5010	5000	100	90 - 110	P	08/20/2025	14:39	LB136898
	Silver	1250	1250	100	90 - 110	P	08/20/2025	14:39	LB136898
	Sodium	25200	25000	101	90 - 110	P	08/20/2025	14:39	LB136898
	Thallium	5050	5000	101	90 - 110	P	08/20/2025	14:39	LB136898
	Vanadium	2460	2500	98	90 - 110	P	08/20/2025	14:39	LB136898
	Zinc	2510	2500	100	90 - 110	P	08/20/2025	14:39	LB136898
CCV03	Aluminum	9770	10000	98	90 - 110	P	08/20/2025	15:45	LB136898
	Antimony	5030	5000	101	90 - 110	P	08/20/2025	15:45	LB136898
	Arsenic	5010	5000	100	90 - 110	P	08/20/2025	15:45	LB136898
	Barium	9560	10000	96	90 - 110	P	08/20/2025	15:45	LB136898
	Beryllium	251	250	100	90 - 110	P	08/20/2025	15:45	LB136898
	Cadmium	2470	2500	99	90 - 110	P	08/20/2025	15:45	LB136898
	Calcium	24100	25000	96	90 - 110	P	08/20/2025	15:45	LB136898
	Chromium	1020	1000	102	90 - 110	P	08/20/2025	15:45	LB136898
	Cobalt	2460	2500	98	90 - 110	P	08/20/2025	15:45	LB136898
	Copper	1250	1250	100	90 - 110	P	08/20/2025	15:45	LB136898
	Iron	4950	5000	99	90 - 110	P	08/20/2025	15:45	LB136898
	Lead	4900	5000	98	90 - 110	P	08/20/2025	15:45	LB136898
	Magnesium	24100	25000	96	90 - 110	P	08/20/2025	15:45	LB136898
	Manganese	2390	2500	96	90 - 110	P	08/20/2025	15:45	LB136898
	Nickel	2470	2500	99	90 - 110	P	08/20/2025	15:45	LB136898
	Potassium	24500	25000	98	90 - 110	P	08/20/2025	15:45	LB136898
	Selenium	5050	5000	101	90 - 110	P	08/20/2025	15:45	LB136898
	Silver	1260	1250	101	90 - 110	P	08/20/2025	15:45	LB136898
	Sodium	22900	25000	92	90 - 110	P	08/20/2025	15:45	LB136898
	Thallium	4780	5000	96	90 - 110	P	08/20/2025	15:45	LB136898
	Vanadium	2440	2500	98	90 - 110	P	08/20/2025	15:45	LB136898
	Zinc	2530	2500	101	90 - 110	P	08/20/2025	15:45	LB136898
CCV04	Aluminum	9870	10000	99	90 - 110	P	08/20/2025	16:35	LB136898
	Antimony	5050	5000	101	90 - 110	P	08/20/2025	16:35	LB136898

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: BAPS North Bergen Development Inc.

SDG No.: Q2892

Contract: BAPS03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	5050	5000	101	90 - 110	P	08/20/2025	16:35	LB136898
	Barium	9770	10000	98	90 - 110	P	08/20/2025	16:35	LB136898
	Beryllium	247	250	99	90 - 110	P	08/20/2025	16:35	LB136898
	Cadmium	2480	2500	99	90 - 110	P	08/20/2025	16:35	LB136898
	Calcium	24600	25000	98	90 - 110	P	08/20/2025	16:35	LB136898
	Chromium	1010	1000	101	90 - 110	P	08/20/2025	16:35	LB136898
	Cobalt	2480	2500	99	90 - 110	P	08/20/2025	16:35	LB136898
	Copper	1260	1250	100	90 - 110	P	08/20/2025	16:35	LB136898
	Iron	4980	5000	100	90 - 110	P	08/20/2025	16:35	LB136898
	Lead	4930	5000	99	90 - 110	P	08/20/2025	16:35	LB136898
	Magnesium	24500	25000	98	90 - 110	P	08/20/2025	16:35	LB136898
	Manganese	2430	2500	97	90 - 110	P	08/20/2025	16:35	LB136898
	Nickel	2480	2500	99	90 - 110	P	08/20/2025	16:35	LB136898
	Potassium	24800	25000	99	90 - 110	P	08/20/2025	16:35	LB136898
	Selenium	5090	5000	102	90 - 110	P	08/20/2025	16:35	LB136898
	Silver	1250	1250	100	90 - 110	P	08/20/2025	16:35	LB136898
	Sodium	23800	25000	95	90 - 110	P	08/20/2025	16:35	LB136898
	Thallium	5150	5000	103	90 - 110	P	08/20/2025	16:35	LB136898
	Vanadium	2470	2500	99	90 - 110	P	08/20/2025	16:35	LB136898
	Zinc	2530	2500	101	90 - 110	P	08/20/2025	16:35	LB136898
CCV05	Aluminum	9860	10000	99	90 - 110	P	08/20/2025	17:09	LB136898
	Antimony	4990	5000	100	90 - 110	P	08/20/2025	17:09	LB136898
	Arsenic	4990	5000	100	90 - 110	P	08/20/2025	17:09	LB136898
	Barium	9690	10000	97	90 - 110	P	08/20/2025	17:09	LB136898
	Beryllium	251	250	101	90 - 110	P	08/20/2025	17:09	LB136898
	Cadmium	2480	2500	99	90 - 110	P	08/20/2025	17:09	LB136898
	Calcium	24500	25000	98	90 - 110	P	08/20/2025	17:09	LB136898
	Chromium	1010	1000	102	90 - 110	P	08/20/2025	17:09	LB136898
	Cobalt	2470	2500	99	90 - 110	P	08/20/2025	17:09	LB136898
	Copper	1250	1250	100	90 - 110	P	08/20/2025	17:09	LB136898
	Iron	4930	5000	98	90 - 110	P	08/20/2025	17:09	LB136898
	Lead	4920	5000	98	90 - 110	P	08/20/2025	17:09	LB136898
	Magnesium	24500	25000	98	90 - 110	P	08/20/2025	17:09	LB136898
	Manganese	2420	2500	97	90 - 110	P	08/20/2025	17:09	LB136898

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: BAPS North Bergen Development Inc.

SDG No.: Q2892

Contract: BAPS03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2480	2500	99	90 - 110	P	08/20/2025	17:09	LB136898
	Potassium	24600	25000	98	90 - 110	P	08/20/2025	17:09	LB136898
	Selenium	5020	5000	100	90 - 110	P	08/20/2025	17:09	LB136898
	Silver	1250	1250	100	90 - 110	P	08/20/2025	17:09	LB136898
	Sodium	23400	25000	94	90 - 110	P	08/20/2025	17:09	LB136898
	Thallium	5030	5000	101	90 - 110	P	08/20/2025	17:09	LB136898
	Vanadium	2460	2500	99	90 - 110	P	08/20/2025	17:09	LB136898
	Zinc	2540	2500	102	90 - 110	P	08/20/2025	17:09	LB136898