

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: BAPS North Bergen Development Inc.

SDG No.: Q2892

Contract: BAPS03

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169292BS Mercury	mg/Kg	0.28	0.23		84	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: BAPS North Bergen Development Inc.

SDG No.: Q2892

Contract: BAPS03

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169304BS							
Aluminum	mg/Kg	100	94.0		94	80 - 120	P
Antimony	mg/Kg	40.0	38.4		96	80 - 120	P
Arsenic	mg/Kg	40.0	37.7		94	80 - 120	P
Barium	mg/Kg	10.0	8.87		89	80 - 120	P
Beryllium	mg/Kg	10.0	9.28		93	80 - 120	P
Cadmium	mg/Kg	10.0	9.01		90	80 - 120	P
Calcium	mg/Kg	50.0	48.1	J	96	80 - 120	P
Chromium	mg/Kg	20.0	19.1		96	80 - 120	P
Cobalt	mg/Kg	10.0	9.23		92	80 - 120	P
Copper	mg/Kg	15.0	14.8		99	80 - 120	P
Iron	mg/Kg	150	144		96	80 - 120	P
Lead	mg/Kg	50.0	45.3		91	80 - 120	P
Magnesium	mg/Kg	100	94.1	J	94	80 - 120	P
Manganese	mg/Kg	10.0	9.42		94	80 - 120	P
Nickel	mg/Kg	25.0	23.2		93	80 - 120	P
Potassium	mg/Kg	500	469		94	80 - 120	P
Selenium	mg/Kg	100	96.2		96	80 - 120	P
Silver	mg/Kg	3.8	3.56		94	80 - 120	P
Sodium	mg/Kg	150	153		102	80 - 120	P
Thallium	mg/Kg	100	98.9		99	80 - 120	P
Vanadium	mg/Kg	15.0	14.2		95	80 - 120	P
Zinc	mg/Kg	10.0	9.95		100	80 - 120	P