

Continuing Calibration Report for SequenceID : FC082225AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069716.D

Aliphatic C16-C21	7520538.000	60.000	9.997	13.364	125342.300	133561.231	6.154
Aliphatic C21-C28	8793009.000	80.000	13.365	17.031	109912.613	118307.822	7.096
Aliphatic C9-C12	8250673.000	60.000	3.298	6.594	137511.217	144679.273	4.954
Aliphatic C12-C16	5467283.000	40.000	6.595	9.996	136682.075	143435.206	4.708
Aliphatic C28-C40	10712882.000	120.000	17.032	22.001	89274.017	95315.213	6.338
Aliphatic EPH	40744385.000	360.000	3.298	22.001	113178.847	120373.027	5.977