

Continuing Calibration Report for SequenceID : FC082525AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069719.D

Aliphatic C28-C40	11205611.000	120.000	17.032	22.006	93380.092	95315.213	2.030
Aliphatic C9-C12	8443573.000	60.000	3.297	6.594	140726.217	144679.273	2.732
Aliphatic C12-C16	5712512.000	40.000	6.595	9.997	142812.800	143435.206	0.434
Aliphatic C16-C21	7949654.000	60.000	9.998	13.365	132494.233	133561.231	0.799
Aliphatic C21-C28	9365203.000	80.000	13.366	17.031	117065.038	118307.822	1.050
Aliphatic EPH	42676553.000	360.000	3.297	22.006	118545.981	120373.027	1.518