

Continuing Calibration Report for SequenceID : FC082525AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069726.D

Aliphatic C28-C40	11281235.000	120.000	17.032	22.006	94010.292	95315.213	1.369
Aliphatic C9-C12	8606704.000	60.000	3.297	6.594	143445.067	144679.273	0.853
Aliphatic C12-C16	5831688.000	40.000	6.595	9.997	145792.200	143435.206	-1.643
Aliphatic C16-C21	8074031.000	60.000	9.998	13.365	134567.183	133561.231	-0.753
Aliphatic C21-C28	9451649.000	80.000	13.366	17.031	118145.613	118307.822	0.137
Aliphatic EPH	43245307.000	360.000	3.297	22.006	120125.853	120373.027	0.205