

Continuing Calibration Report for SequenceID : FD082225AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049645.D

Aromatic C10-C12	4751542.000	40.000	4.407	6.150	118788.550	107939.663	-10.051
Aromatic C12-C16	8287038.000	60.000	6.151	8.778	138117.300	123580.152	-11.763
Aromatic C16-C21	10674121.000	80.000	8.779	13.069	133426.513	126486.361	-5.487
Aromatic C21-C36	20574666.000	180.000	13.070	18.489	114303.700	112618.254	-1.497
Aromatic EPH	44287367.000	360.000	4.407	18.489	123020.464	117007.195	-5.139