

Continuing Calibration Report for SequenceID : FD082225AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049655.D

Aromatic C10-C12	4855203.000	40.000	4.407	6.150	121380.075	107939.663	-12.452
Aromatic C12-C16	8282990.000	60.000	6.151	8.778	138049.833	123580.152	-11.709
Aromatic C16-C21	10768702.000	80.000	8.779	13.069	134608.775	126486.361	-6.422
Aromatic C21-C36	20980863.000	180.000	13.070	18.489	116560.350	112618.254	-3.500
Aromatic EPH	44887758.000	360.000	4.407	18.489	124688.217	117007.195	-6.565