

Continuing Calibration Report for SequenceID : FD082225AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049666.D

Aromatic C10-C12	4410803.000	40.000	4.407	6.150	110270.075	107939.663	-2.159
Aromatic C12-C16	7531907.000	60.000	6.151	8.778	125531.783	123580.152	-1.579
Aromatic C16-C21	10146229.000	80.000	8.779	13.069	126827.863	126486.361	-0.270
Aromatic C21-C36	21272022.000	180.000	13.070	18.489	118177.900	112618.254	-4.937
Aromatic EPH	43360961.000	360.000	4.407	18.489	120447.114	117007.195	-2.940