

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** BAPS03  
**Lab Code:** ACE **SDG NO.:** Q2892  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

**Client Sample No.:** CCAL01 **Date Analyzed:** 08/19/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO113038.D **Time Analyzed:** 08:41

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.747 | 4.648     | 4.848 | 525.350            | 500.000           | 5.1  |
| Aroclor-1016-2       | 4.765 | 4.666     | 4.866 | 535.020            | 500.000           | 7.0  |
| Aroclor-1016-3       | 4.823 | 4.724     | 4.924 | 537.240            | 500.000           | 7.4  |
| Aroclor-1016-4       | 4.941 | 4.843     | 5.043 | 488.440            | 500.000           | -2.3 |
| Aroclor-1016-5       | 5.199 | 5.099     | 5.299 | 520.510            | 500.000           | 4.1  |
| Aroclor-1260-1       | 6.235 | 6.136     | 6.336 | 514.780            | 500.000           | 3.0  |
| Aroclor-1260-2       | 6.425 | 6.326     | 6.526 | 508.940            | 500.000           | 1.8  |
| Aroclor-1260-3       | 6.789 | 6.691     | 6.891 | 519.130            | 500.000           | 3.8  |
| Aroclor-1260-4       | 7.049 | 6.951     | 7.151 | 526.330            | 500.000           | 5.3  |
| Aroclor-1260-5       | 7.293 | 7.194     | 7.394 | 532.510            | 500.000           | 6.5  |
| Decachlorobiphenyl   | 8.681 | 8.582     | 8.782 | 52.810             | 50.000            | 5.6  |
| Tetrachloro-m-xylene | 3.662 | 3.563     | 3.763 | 52.690             | 50.000            | 5.4  |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** BAPS03  
**Lab Code:** ACE **SDG NO.:** Q2892  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

**Client Sample No.:** CCAL01 **Date Analyzed:** 08/19/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO113038.D **Time Analyzed:** 08:41

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.728 | 4.629     | 4.829 | 507.340            | 500.000           | 1.5  |
| Aroclor-1016-2       | 4.746 | 4.647     | 4.847 | 503.570            | 500.000           | 0.7  |
| Aroclor-1016-3       | 4.921 | 4.822     | 5.022 | 490.950            | 500.000           | -1.8 |
| Aroclor-1016-4       | 4.963 | 4.864     | 5.064 | 477.490            | 500.000           | -4.5 |
| Aroclor-1016-5       | 5.175 | 5.076     | 5.276 | 507.500            | 500.000           | 1.5  |
| Aroclor-1260-1       | 6.202 | 6.104     | 6.304 | 486.450            | 500.000           | -2.7 |
| Aroclor-1260-2       | 6.391 | 6.292     | 6.492 | 484.140            | 500.000           | -3.2 |
| Aroclor-1260-3       | 6.542 | 6.443     | 6.643 | 470.190            | 500.000           | -6.0 |
| Aroclor-1260-4       | 7.011 | 6.912     | 7.112 | 484.800            | 500.000           | -3.0 |
| Aroclor-1260-5       | 7.254 | 7.155     | 7.355 | 475.930            | 500.000           | -4.8 |
| Decachlorobiphenyl   | 8.625 | 8.528     | 8.728 | 50.770             | 50.000            | 1.5  |
| Tetrachloro-m-xylene | 3.654 | 3.555     | 3.755 | 51.110             | 50.000            | 2.2  |