

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** BAPS03
Lab Code: ACE **SDG NO.:** Q2892
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/19/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113053.D **Time Analyzed:** 15:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.749	4.648	4.848	524.080	500.000	4.8
Aroclor-1016-2	4.767	4.666	4.866	517.550	500.000	3.5
Aroclor-1016-3	4.824	4.724	4.924	525.380	500.000	5.1
Aroclor-1016-4	4.944	4.843	5.043	483.260	500.000	-3.3
Aroclor-1016-5	5.200	5.099	5.299	535.500	500.000	7.1
Aroclor-1260-1	6.236	6.136	6.336	504.100	500.000	0.8
Aroclor-1260-2	6.426	6.326	6.526	505.550	500.000	1.1
Aroclor-1260-3	6.791	6.691	6.891	512.140	500.000	2.4
Aroclor-1260-4	7.050	6.951	7.151	509.950	500.000	2.0
Aroclor-1260-5	7.293	7.194	7.394	514.330	500.000	2.9
Decachlorobiphenyl	8.682	8.582	8.782	51.600	50.000	3.2
Tetrachloro-m-xylene	3.664	3.563	3.763	50.950	50.000	1.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** BAPS03
Lab Code: ACE **SDG NO.:** Q2892
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/19/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113053.D **Time Analyzed:** 15:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.729	4.629	4.829	485.530	500.000	-2.9
Aroclor-1016-2	4.748	4.647	4.847	491.270	500.000	-1.7
Aroclor-1016-3	4.922	4.822	5.022	476.600	500.000	-4.7
Aroclor-1016-4	4.965	4.864	5.064	463.350	500.000	-7.3
Aroclor-1016-5	5.177	5.076	5.276	499.450	500.000	-0.1
Aroclor-1260-1	6.203	6.104	6.304	473.780	500.000	-5.2
Aroclor-1260-2	6.393	6.292	6.492	472.110	500.000	-5.6
Aroclor-1260-3	6.543	6.443	6.643	447.990	500.000	-10.4
Aroclor-1260-4	7.013	6.912	7.112	465.640	500.000	-6.9
Aroclor-1260-5	7.255	7.155	7.355	458.280	500.000	-8.3
Decachlorobiphenyl	8.627	8.528	8.728	48.620	50.000	-2.8
Tetrachloro-m-xylene	3.656	3.555	3.755	49.380	50.000	-1.2