

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** BAPS03
Lab Code: ACE **SDG NO.:** Q2892
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/19/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113068.D **Time Analyzed:** 21:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.747	4.648	4.848	566.570	500.000	13.3
Aroclor-1016-2	4.766	4.666	4.866	572.480	500.000	14.5
Aroclor-1016-3	4.823	4.724	4.924	568.570	500.000	13.7
Aroclor-1016-4	4.942	4.843	5.043	523.290	500.000	4.7
Aroclor-1016-5	5.198	5.099	5.299	559.260	500.000	11.9
Aroclor-1260-1	6.234	6.136	6.336	450.840	500.000	-9.8
Aroclor-1260-2	6.425	6.326	6.526	509.860	500.000	2.0
Aroclor-1260-3	6.790	6.691	6.891	461.610	500.000	-7.7
Aroclor-1260-4	7.049	6.951	7.151	414.460	500.000	-17.1
Aroclor-1260-5	7.293	7.194	7.394	407.460	500.000	-18.5
Decachlorobiphenyl	8.681	8.582	8.782	43.080	50.000	-13.8
Tetrachloro-m-xylene	3.664	3.563	3.763	56.420	50.000	12.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** BAPS03
Lab Code: ACE **SDG NO.:** Q2892
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/19/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113068.D **Time Analyzed:** 21:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.729	4.629	4.829	535.260	500.000	7.1
Aroclor-1016-2	4.747	4.647	4.847	528.350	500.000	5.7
Aroclor-1016-3	4.921	4.822	5.022	518.520	500.000	3.7
Aroclor-1016-4	4.964	4.864	5.064	499.520	500.000	-0.1
Aroclor-1016-5	5.175	5.076	5.276	558.700	500.000	11.7
Aroclor-1260-1	6.203	6.104	6.304	459.530	500.000	-8.1
Aroclor-1260-2	6.391	6.292	6.492	462.080	500.000	-7.6
Aroclor-1260-3	6.543	6.443	6.643	431.220	500.000	-13.8
Aroclor-1260-4	7.012	6.912	7.112	426.240	500.000	-14.8
Aroclor-1260-5	7.254	7.155	7.355	421.820	500.000	-15.6
Decachlorobiphenyl	8.627	8.528	8.728	46.740	50.000	-6.5
Tetrachloro-m-xylene	3.656	3.555	3.755	53.750	50.000	7.5