

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO081825

Review By	yogesh	Review On	8/18/2025 10:23:59 AM
Supervise By	mohammad	Supervise On	8/20/2025 3:04:50 AM
SubDirectory	PO081825	HP Acquire Method	HP Processing Method PO081825

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP2416,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO113001.D	18 Aug 2025 09:16		YP/AJ	Ok
2	I.BLK	I.BLK	PO113002.D	18 Aug 2025 09:34		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO113003.D	18 Aug 2025 09:53		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO113004.D	18 Aug 2025 10:11		YP/AJ	Ok
5	AR1660ICC500	AR1660ICC500	PO113005.D	18 Aug 2025 10:29		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO113006.D	18 Aug 2025 10:48		YP/AJ	Ok
7	AR1660ICC050	AR1660ICC050	PO113007.D	18 Aug 2025 11:06		YP/AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PO113008.D	18 Aug 2025 11:25		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO113009.D	18 Aug 2025 11:43		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO113010.D	18 Aug 2025 12:01		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO113011.D	18 Aug 2025 12:20		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO113012.D	18 Aug 2025 12:38		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO113013.D	18 Aug 2025 12:57		YP/AJ	Ok
14	AR1242ICC050	AR1242ICC050	PO113014.D	18 Aug 2025 13:15		YP/AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PO113015.D	18 Aug 2025 13:34		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO113016.D	18 Aug 2025 13:52		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO113017.D	18 Aug 2025 14:09		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO113018.D	18 Aug 2025 14:28		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO081825

Review By	yogesh	Review On	8/18/2025 10:23:59 AM
Supervise By	mohammad	Supervise On	8/20/2025 3:04:50 AM
SubDirectory	PO081825	HP Acquire Method	HP Processing Method PO081825
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP2416,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

19	AR1248ICC050	AR1248ICC050	PO113019.D	18 Aug 2025 14:46		YP/AJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PO113020.D	18 Aug 2025 15:05		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO113021.D	18 Aug 2025 15:23		YP/AJ	Ok
22	AR1254ICC500	AR1254ICC500	PO113022.D	18 Aug 2025 15:41		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO113023.D	18 Aug 2025 16:00		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO113024.D	18 Aug 2025 16:18		YP/AJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PO113025.D	18 Aug 2025 16:54		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO113026.D	18 Aug 2025 17:12		YP/AJ	Ok
27	AR1268ICC750	AR1268ICC750	PO113027.D	18 Aug 2025 17:30		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO113028.D	18 Aug 2025 17:48		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO113029.D	18 Aug 2025 18:06		YP/AJ	Ok
30	AR1268ICC050	AR1268ICC050	PO113030.D	18 Aug 2025 18:24		YP/AJ	Ok,M
31	PO081825ICV500	ICVPO081825	PO113031.D	18 Aug 2025 18:43		YP/AJ	Ok
32	AR1242ICV500	ICVPO081825AR1242	PO113032.D	18 Aug 2025 19:37		YP/AJ	Ok
33	AR1248ICV500	ICVPO081825AR1248	PO113033.D	18 Aug 2025 19:54		YP/AJ	Ok
34	AR1254ICV500	ICVPO081825AR1254	PO113034.D	18 Aug 2025 20:31		YP/AJ	Ok
35	AR1268ICV500	ICVPO081825AR1268	PO113035.D	18 Aug 2025 21:07		YP/AJ	Ok,M
36	DDT ANALOG	DDT ANALOG	PO113036.D	18 Aug 2025 21:44		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO081925

Review By	yogesh	Review On	8/19/2025 10:42:44 AM
Supervise By		Supervise On	
SubDirectory	PO081925	HP Acquire Method	HP Processing Method PO081825

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP2416,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO113037.D	19 Aug 2025 08:23		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO113038.D	19 Aug 2025 08:41		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO113039.D	19 Aug 2025 09:02		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO113040.D	19 Aug 2025 09:20		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO113041.D	19 Aug 2025 09:39		YP/AJ	Ok
6	I.BLK	I.BLK	PO113042.D	19 Aug 2025 09:57		YP/AJ	Ok
7	DDT ANALOG	DDT ANALOG	PO113043.D	19 Aug 2025 10:16		YP/AJ	Ok
8	Q2865-03	LAW-25-0109	PO113044.D	19 Aug 2025 10:34	TCMX LOW in both column	YP/AJ	ReRun
9	PB169297BL	PB169297BL	PO113045.D	19 Aug 2025 12:22		YP/AJ	Ok
10	PB169297BS	PB169297BS	PO113046.D	19 Aug 2025 12:40		YP/AJ	Ok
11	Q2895-01	1104	PO113047.D	19 Aug 2025 12:58		YP/AJ	Ok,NS
12	Q2895-03	1204	PO113048.D	19 Aug 2025 13:17	AR1254 Hit	YP/AJ	Ok
13	Q2891-01	BARRICADE	PO113049.D	19 Aug 2025 13:35		YP/AJ	Ok
14	Q2891-02	WALL-1	PO113050.D	19 Aug 2025 13:54		YP/AJ	Ok
15	Q2891-03	WALL-2	PO113051.D	19 Aug 2025 14:12		YP/AJ	Ok
16	Q2891-04	WALL-3	PO113052.D	19 Aug 2025 14:30		YP/AJ	Ok
17	AR1660CCC500	AR1660CCC500	PO113053.D	19 Aug 2025 15:43		YP/AJ	Ok
18	AR1242CCC500	AR1242CCC500	PO113054.D	19 Aug 2025 16:01		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO081925

Review By	yogesh	Review On	8/19/2025 10:42:44 AM
Supervise By		Supervise On	
SubDirectory	PO081925	HP Acquire Method	HP Processing Method PO081825
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP2416,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PO113055.D	19 Aug 2025 16:20		YP/AJ	Ok
20	AR1254CCC500	AR1254CCC500	PO113056.D	19 Aug 2025 16:38		YP/AJ	Ok
21	I.BLK	I.BLK	PO113057.D	19 Aug 2025 16:57		YP/AJ	Ok
22	Q2892-02	1-FLOOR-NORTH-EAS	PO113058.D	19 Aug 2025 17:14		YP/AJ	Ok
23	Q2892-04	1-FLOOR-DRILL-B-BA	PO113059.D	19 Aug 2025 17:32	AR1260 Hit	YP/AJ	Ok,NR
24	Q2892-06	BACK-1-FLOOR-CENT	PO113060.D	19 Aug 2025 17:51		YP/AJ	Ok
25	Q2892-08	3-FLOOR-NORTH-D	PO113061.D	19 Aug 2025 18:09		YP/AJ	Ok
26	Q2892-10	3-FLOOR-SOUTH-E	PO113062.D	19 Aug 2025 18:27		YP/AJ	Ok
27	Q2892-12	FRONT-1-FLOOR-F	PO113063.D	19 Aug 2025 18:46		YP/AJ	Ok
28	Q2897-01	HD-02-08182025	PO113064.D	19 Aug 2025 19:04		YP/AJ	Ok
29	Q2897-01MS	HD-02-08182025MS	PO113065.D	19 Aug 2025 19:23		YP/AJ	Ok
30	Q2897-01MSD	HD-02-08182025MSD	PO113066.D	19 Aug 2025 19:41		YP/AJ	Ok,NS
31	Q2865-03RE	LAW-25-0109	PO113067.D	19 Aug 2025 19:58	TCMX LOW in both column	YP/AJ	Confirms
32	AR1660CCC500	AR1660CCC500	PO113068.D	19 Aug 2025 21:30		YP/AJ	Ok,NS
33	AR1242CCC500	AR1242CCC500	PO113069.D	19 Aug 2025 21:49		YP/AJ	Ok,NS
34	AR1248CCC500	AR1248CCC500	PO113070.D	19 Aug 2025 22:07		YP/AJ	Ok
35	AR1254CCC500	AR1254CCC500	PO113071.D	19 Aug 2025 22:26		YP/AJ	Ok
36	I.BLK	I.BLK	PO113072.D	19 Aug 2025 22:43		YP/AJ	Ok

M : Manual Integration