

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** BAPS03
Lab Code: ACE **SDG NO.:** Q2892
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/28/2025 07/28/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/19/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096905.D **Time Analyzed:** 18:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.667	6.575	6.775	57.840	50.000	15.7
4,4'-DDE	6.159	6.066	6.266	54.220	50.000	8.4
4,4'-DDT	6.982	6.889	7.089	52.300	50.000	4.6
Aldrin	5.236	5.143	5.343	53.760	50.000	7.5
alpha-BHC	3.976	3.882	4.082	55.300	50.000	10.6
alpha-Chlordane	5.989	5.897	6.097	53.370	50.000	6.7
beta-BHC	4.490	4.397	4.597	55.970	50.000	11.9
Decachlorobiphenyl	9.003	8.916	9.116	53.530	50.000	7.1
delta-BHC	4.737	4.643	4.843	56.020	50.000	12.0
Dieldrin	6.307	6.216	6.416	52.460	50.000	4.9
Endosulfan I	6.037	5.944	6.144	52.240	50.000	4.5
Endosulfan II	6.746	6.655	6.855	51.610	50.000	3.2
Endosulfan sulfate	7.108	7.018	7.218	50.840	50.000	1.7
Endrin	6.533	6.442	6.642	53.090	50.000	6.2
Endrin aldehyde	6.876	6.784	6.984	54.570	50.000	9.1
Endrin ketone	7.589	7.497	7.697	50.160	50.000	0.3
gamma-BHC (Lindane)	4.304	4.210	4.410	54.790	50.000	9.6
gamma-Chlordane	5.909	5.816	6.016	54.380	50.000	8.8
Heptachlor	4.896	4.803	5.003	57.040	50.000	14.1
Heptachlor epoxide	5.654	5.562	5.762	55.200	50.000	10.4
Methoxychlor	7.454	7.362	7.562	50.370	50.000	0.7
Tetrachloro-m-xylene	3.531	3.435	3.635	54.090	50.000	8.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** BAPS03
Lab Code: ACE **SDG NO.:** Q2892
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/28/2025 07/28/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/19/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096905.D **Time Analyzed:** 18:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.852	5.756	5.956	51.980	50.000	4.0
4,4'-DDE	5.299	5.204	5.404	52.450	50.000	4.9
4,4'-DDT	6.104	6.009	6.209	50.190	50.000	0.4
Aldrin	4.293	4.197	4.397	52.910	50.000	5.8
alpha-BHC	3.331	3.233	3.433	53.360	50.000	6.7
alpha-Chlordane	5.111	5.015	5.215	51.930	50.000	3.9
beta-BHC	3.958	3.861	4.061	53.200	50.000	6.4
Decachlorobiphenyl	7.986	7.892	8.092	48.960	50.000	-2.1
delta-BHC	4.191	4.095	4.295	53.230	50.000	6.5
Dieldrin	5.429	5.334	5.534	52.300	50.000	4.6
Endosulfan I	5.165	5.069	5.269	49.080	50.000	-1.8
Endosulfan II	5.996	5.901	6.101	50.960	50.000	1.9
Endosulfan sulfate	6.397	6.302	6.502	49.950	50.000	-0.1
Endrin	5.705	5.609	5.809	55.460	50.000	10.9
Endrin aldehyde	6.174	6.079	6.279	52.070	50.000	4.1
Endrin ketone	6.901	6.807	7.007	52.290	50.000	4.6
gamma-BHC (Lindane)	3.663	3.565	3.765	53.070	50.000	6.1
gamma-Chlordane	5.047	4.951	5.151	52.950	50.000	5.9
Heptachlor	4.011	3.914	4.114	52.180	50.000	4.4
Heptachlor epoxide	4.795	4.699	4.899	52.020	50.000	4.0
Methoxychlor	6.675	6.581	6.781	49.970	50.000	-0.1
Tetrachloro-m-xylene	2.826	2.728	2.928	52.880	50.000	5.8