

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** BAPS03
Lab Code: ACE **SDG NO.:** Q2892
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/28/2025 07/28/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/19/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096916.D **Time Analyzed:** 22:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.668	6.575	6.775	58.570	50.000	17.1
4,4'-DDE	6.159	6.066	6.266	54.650	50.000	9.3
4,4'-DDT	6.982	6.889	7.089	50.370	50.000	0.7
Aldrin	5.236	5.143	5.343	53.610	50.000	7.2
alpha-BHC	3.977	3.882	4.082	56.000	50.000	12.0
alpha-Chlordane	5.989	5.897	6.097	54.810	50.000	9.6
beta-BHC	4.491	4.397	4.597	57.140	50.000	14.3
Decachlorobiphenyl	9.003	8.916	9.116	53.440	50.000	6.9
delta-BHC	4.737	4.643	4.843	56.630	50.000	13.3
Dieldrin	6.307	6.216	6.416	52.520	50.000	5.0
Endosulfan I	6.036	5.944	6.144	53.670	50.000	7.3
Endosulfan II	6.746	6.655	6.855	50.830	50.000	1.7
Endosulfan sulfate	7.109	7.018	7.218	50.700	50.000	1.4
Endrin	6.533	6.442	6.642	51.720	50.000	3.4
Endrin aldehyde	6.877	6.784	6.984	54.540	50.000	9.1
Endrin ketone	7.589	7.497	7.697	49.730	50.000	-0.5
gamma-BHC (Lindane)	4.305	4.210	4.410	55.010	50.000	10.0
gamma-Chlordane	5.908	5.816	6.016	57.240	50.000	14.5
Heptachlor	4.896	4.803	5.003	57.140	50.000	14.3
Heptachlor epoxide	5.653	5.562	5.762	54.670	50.000	9.3
Methoxychlor	7.454	7.362	7.562	48.850	50.000	-2.3
Tetrachloro-m-xylene	3.531	3.435	3.635	55.440	50.000	10.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** BAPS03
Lab Code: ACE **SDG NO.:** Q2892
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/28/2025 07/28/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/19/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096916.D **Time Analyzed:** 22:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.852	5.756	5.956	49.990	50.000	0.0
4,4'-DDE	5.300	5.204	5.404	50.880	50.000	1.8
4,4'-DDT	6.104	6.009	6.209	45.840	50.000	-8.3
Aldrin	4.293	4.197	4.397	53.080	50.000	6.2
alpha-BHC	3.331	3.233	3.433	53.480	50.000	7.0
alpha-Chlordane	5.111	5.015	5.215	50.580	50.000	1.2
beta-BHC	3.959	3.861	4.061	53.140	50.000	6.3
Decachlorobiphenyl	7.985	7.892	8.092	43.420	50.000	-13.2
delta-BHC	4.191	4.095	4.295	52.880	50.000	5.8
Dieldrin	5.429	5.334	5.534	50.890	50.000	1.8
Endosulfan I	5.166	5.069	5.269	47.690	50.000	-4.6
Endosulfan II	5.996	5.901	6.101	48.360	50.000	-3.3
Endosulfan sulfate	6.398	6.302	6.502	46.490	50.000	-7.0
Endrin	5.705	5.609	5.809	52.630	50.000	5.3
Endrin aldehyde	6.175	6.079	6.279	49.330	50.000	-1.3
Endrin ketone	6.902	6.807	7.007	48.880	50.000	-2.2
gamma-BHC (Lindane)	3.663	3.565	3.765	53.140	50.000	6.3
gamma-Chlordane	5.047	4.951	5.151	51.820	50.000	3.6
Heptachlor	4.011	3.914	4.114	51.930	50.000	3.9
Heptachlor epoxide	4.795	4.699	4.899	51.360	50.000	2.7
Methoxychlor	6.675	6.581	6.781	45.740	50.000	-8.5
Tetrachloro-m-xylene	2.826	2.728	2.928	53.000	50.000	6.0