

Report of Analysis

| | | | | | |
|--------------------|------------------------------------|-----------|--------------------|---------------|-----------|
| Client: | BAPS North Bergen Development Inc. | | Date Collected: | 08/19/25 | |
| Project: | BAPS 2000 Tonnelles Ave | | Date Received: | 08/19/25 | |
| Client Sample ID: | PIBLK-PL096915.D | | SDG No.: | Q2892 | |
| Lab Sample ID: | I.BLK-PL096915.D | | Matrix: | WATER | |
| Analytical Method: | 8081B | | % Solid: | 0 | Decanted: |
| Sample Wt/Vol: | 1000 | Units: mL | Final Vol: | 10000 | uL |
| Soil Aliquot Vol: | | uL | Test: | Pesticide-TCL | |
| Extraction Type: | | | Injection Volume : | | |
| GPC Factor : | 1.0 | PH : | | | |
| Prep Method : | 3510C | | | | |

| | | | | |
|-------------------|-----------|-----------|---------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
| PL096915.D | 1 | | 08/19/25 | pl081925 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|-------------------|----------------------|--------|-----------|---------------------|------------|---------|
| TARGETS | | | | | | |
| 319-84-6 | alpha-BHC | 0.0039 | U | 0.0039 | 0.050 | ug/L |
| 319-85-7 | beta-BHC | 0.0049 | U | 0.0049 | 0.050 | ug/L |
| 319-86-8 | delta-BHC | 0.011 | U | 0.011 | 0.050 | ug/L |
| 58-89-9 | gamma-BHC (Lindane) | 0.0037 | U | 0.0037 | 0.050 | ug/L |
| 76-44-8 | Heptachlor | 0.0027 | U | 0.0027 | 0.050 | ug/L |
| 309-00-2 | Aldrin | 0.0036 | U | 0.0036 | 0.050 | ug/L |
| 1024-57-3 | Heptachlor epoxide | 0.0096 | U | 0.0096 | 0.050 | ug/L |
| 959-98-8 | Endosulfan I | 0.0031 | U | 0.0031 | 0.050 | ug/L |
| 60-57-1 | Dieldrin | 0.0036 | U | 0.0036 | 0.050 | ug/L |
| 72-55-9 | 4,4-DDE | 0.0037 | U | 0.0037 | 0.050 | ug/L |
| 72-20-8 | Endrin | 0.0032 | U | 0.0032 | 0.050 | ug/L |
| 33213-65-9 | Endosulfan II | 0.0079 | U | 0.0079 | 0.050 | ug/L |
| 72-54-8 | 4,4-DDD | 0.0071 | U | 0.0071 | 0.050 | ug/L |
| 1031-07-8 | Endosulfan Sulfate | 0.0037 | U | 0.0037 | 0.050 | ug/L |
| 50-29-3 | 4,4-DDT | 0.0035 | U | 0.0035 | 0.050 | ug/L |
| 72-43-5 | Methoxychlor | 0.011 | U | 0.011 | 0.050 | ug/L |
| 53494-70-5 | Endrin ketone | 0.0093 | U | 0.0093 | 0.050 | ug/L |
| 7421-93-4 | Endrin aldehyde | 0.011 | U | 0.011 | 0.050 | ug/L |
| 5103-71-9 | alpha-Chlordane | 0.0035 | U | 0.0035 | 0.050 | ug/L |
| 5103-74-2 | gamma-Chlordane | 0.0039 | U | 0.0039 | 0.050 | ug/L |
| 8001-35-2 | Toxaphene | 0.17 | U | 0.17 | 1.00 | ug/L |
| SURROGATES | | | | | | |
| 2051-24-3 | Decachlorobiphenyl | 24.0 | | 30 (57) - 150 (171) | 120% | SPK: 20 |
| 877-09-8 | Tetrachloro-m-xylene | 22.6 | | 30 (61) - 150 (148) | 113% | SPK: 20 |

Report of Analysis

| | | | | | |
|--------------------|------------------------------------|-----------|--------------------|------------------|----|
| Client: | BAPS North Bergen Development Inc. | | Date Collected: | 08/19/25 | |
| Project: | BAPS 2000 Tonnelles Ave | | Date Received: | 08/19/25 | |
| Client Sample ID: | PIBLK-PL096915.D | | SDG No.: | Q2892 | |
| Lab Sample ID: | I.BLK-PL096915.D | | Matrix: | WATER | |
| Analytical Method: | 8081B | | % Solid: | 0 Decanted: | |
| Sample Wt/Vol: | 1000 | Units: mL | Final Vol: | 10000 | uL |
| Soil Aliquot Vol: | | uL | Test: | Pesticide-TCL | |
| Extraction Type: | | | Injection Volume : | | |
| GPC Factor : | 1.0 | PH : | | | |
| Prep Method : | 3510C | | | | |

| | | | | |
|-------------------|-----------|-----------|---------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
| PL096915.D | 1 | | 08/19/25 | p1081925 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit