

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2892</u>	Analytical Method:	<u>8081B</u>
Client:	<u>BAPS North Bergen Development Inc.</u>	DataFile :	<u>PL096911.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits	
			Result								High	RPD
Lab Sample ID:	Q2892-08MSD (Column 1)	Client Sample ID:		3-FLOOR-NORTH-DMSD								
	alpha-BHC	17.69	0	19.0	ug/kg	107		1		30 (60)	150 (144)	30 (20)
	beta-BHC	17.69	0	19.6	ug/kg	111		1		30 (54)	150 (143)	30 (20)
	delta-BHC	17.69	0	15.9	ug/kg	90		0		30 (29)	150 (151)	30 (20)
	gamma-BHC (Lindane)	17.69	0	19.0	ug/kg	107		1		30 (61)	150 (140)	30 (20)
	Heptachlor	17.69	0	19.8	ug/kg	112		1		30 (63)	150 (135)	30 (20)
	Aldrin	17.69	0	18.8	ug/kg	106		0		30 (49)	150 (139)	30 (20)
	Heptachlor epoxide	17.69	0	19.4	ug/kg	110		2		30 (41)	150 (156)	30 (20)
	Endosulfan I	17.69	0	18.6	ug/kg	105		1		30 (56)	150 (142)	30 (20)
	Dieldrin	17.69	0	18.7	ug/kg	106		2		30 (47)	150 (161)	30 (20)
	4,4'-DDE	17.69	0	19.4	ug/kg	110		1		30 (55)	150 (136)	30 (20)
	Endrin	17.69	0	19.0	ug/kg	107		1		30 (57)	150 (139)	30 (20)
	Endosulfan II	17.69	0	18.6	ug/kg	105		1		30 (40)	150 (163)	30 (20)
	4,4'-DDD	17.69	0	19.7	ug/kg	111		1		30 (47)	150 (163)	30 (20)
	Endosulfan sulfate	17.69	0	16.2	ug/kg	92		1		30 (62)	150 (139)	30 (20)
	4,4'-DDT	17.69	0	17.8	ug/kg	101		0		30 (51)	150 (146)	30 (20)
	Methoxychlor	17.69	0	17.0	ug/kg	96		0		30 (54)	150 (136)	30 (20)
	Endrin ketone	17.69	0	17.3	ug/kg	98		0		30 (60)	150 (129)	30 (20)
	Endrin aldehyde	17.69	0	16.2	ug/kg	92		1		30 (59)	150 (132)	30 (20)
	alpha-Chlordane	17.69	0	19.0	ug/kg	107		1		30 (39)	150 (166)	30 (20)
	gamma-Chlordane	17.69	0	19.4	ug/kg	110		1		30 (44)	150 (175)	30 (20)
Lab Sample ID:	Q2892-08MSD (Column 2)	Client Sample ID:		3-FLOOR-NORTH-DMSD								
	alpha-BHC	17.69	0	18.5	ug/kg	105		2		30 (60)	150 (144)	30 (20)
	beta-BHC	17.69	0	18.4	ug/kg	104		1		30 (54)	150 (143)	30 (20)
	delta-BHC	17.69	0	14.9	ug/kg	84		0		30 (29)	150 (151)	30 (20)
	gamma-BHC (Lindane)	17.69	0	18.4	ug/kg	104		2		30 (61)	150 (140)	30 (20)
	Heptachlor	17.69	0	18.0	ug/kg	102		1		30 (63)	150 (135)	30 (20)
	Aldrin	17.69	0	18.4	ug/kg	104		0		30 (49)	150 (139)	30 (20)
	Heptachlor epoxide	17.69	0	18.3	ug/kg	103		0		30 (41)	150 (156)	30 (20)
	Endosulfan I	17.69	0	17.6	ug/kg	99		1		30 (56)	150 (142)	30 (20)
	Dieldrin	17.69	0	18.5	ug/kg	105		0		30 (47)	150 (161)	30 (20)
	4,4'-DDE	17.69	0	18.6	ug/kg	105		0		30 (55)	150 (136)	30 (20)
	Endrin	17.69	0	17.9	ug/kg	101		0		30 (57)	150 (139)	30 (20)
	Endosulfan II	17.69	0	17.5	ug/kg	99		1		30 (40)	150 (163)	30 (20)
	4,4'-DDD	17.69	0	17.8	ug/kg	101		1		30 (47)	150 (163)	30 (20)
	Endosulfan sulfate	17.69	0	15.7	ug/kg	89		2		30 (62)	150 (139)	30 (20)

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2892</u>	Analytical Method:	<u>8081B</u>
Client:	<u>BAPS North Bergen Development Inc.</u>	DataFile :	<u>PL096911.D</u>

Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits	
		Result	Result							High	RPD
4,4'-DDT	17.69	0	17.2	ug/kg	97		1		30 (51)	150 (146)	30 (20)
Methoxychlor	17.69	0	17.0	ug/kg	96		0		30 (54)	150 (136)	30 (20)
Endrin ketone	17.69	0	17.3	ug/kg	98		1		30 (60)	150 (129)	30 (20)
Endrin aldehyde	17.69	0	14.7	ug/kg	83		0		30 (59)	150 (132)	30 (20)
alpha-Chlordane	17.69	0	18.2	ug/kg	103		1		30 (39)	150 (166)	30 (20)
gamma-Chlordane	17.69	0	18.5	ug/kg	105		1		30 (44)	150 (175)	30 (20)