

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: BAPS03
 Lab Code: ACE SDG No.: Q2892
 Instrument ID: BNA_F Calibration Date/Time: 08/22/2025 10:29
 Lab File ID: BF143563.D Init. Calib. Date(s): 08/20/2025 08/20/2025
 EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:55 14:20
 GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.145	1.105		-3.5	
Benzaldehyde	1.145	0.999		-12.8	
Phenol-d6	1.414	1.366		-3.4	
Phenol	1.629	1.599		-1.8	20.0
bis(2-Chloroethyl)ether	1.217	1.157		-4.9	
2-Chlorophenol	1.265	1.220		-3.6	
2-Methylphenol	1.015	0.981		-3.3	
2,2-oxybis(1-Chloropropane)	1.947	1.762		-9.5	
Acetophenone	0.432	0.414		-4.2	
3+4-Methylphenols	1.212	1.227		1.2	
n-Nitroso-di-n-propylamine	0.864	0.857	0.050	-0.8	
Nitrobenzene-d5	0.343	0.333		-2.9	
Hexachloroethane	0.485	0.478		-1.4	
Nitrobenzene	0.353	0.349		-1.1	
Isophorone	0.628	0.612		-2.5	
2-Nitrophenol	0.173	0.169		-2.3	20.0
2,4-Dimethylphenol	0.269	0.253		-5.9	
bis(2-Chloroethoxy)methane	0.386	0.363		-6.0	
2,4-Dichlorophenol	0.288	0.278		-3.5	20.0
Naphthalene	0.960	0.899		-6.4	
4-Chloroaniline	0.341	0.325		-4.7	
Hexachlorobutadiene	0.193	0.186		-3.6	20.0
Caprolactam	0.069	0.078		13.0	
4-Chloro-3-methylphenol	0.288	0.292		1.4	20.0
2-Methylnaphthalene	0.641	0.610		-4.8	
Hexachlorocyclopentadiene	0.184	0.193	0.050	4.9	
2,4,6-Trichlorophenol	0.352	0.336		-4.5	20.0
2-Fluorobiphenyl	1.210	1.143		-5.5	
2,4,5-Trichlorophenol	0.385	0.355		-7.8	
1,1-Biphenyl	1.420	1.319		-7.1	
2-Chloronaphthalene	1.114	1.033		-7.3	
2-Nitroaniline	0.326	0.330		1.2	
Dimethylphthalate	1.205	1.196		-0.7	
Acenaphthylene	1.595	1.486		-6.8	
2,6-Dinitrotoluene	0.250	0.249		-0.4	
3-Nitroaniline	0.270	0.267		-1.1	
Acenaphthene	1.083	1.038		-4.2	20.0
2,4-Dinitrophenol	0.115	0.130	0.050	13.0	
4-Nitrophenol	0.161	0.193	0.050	19.9	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: BAPS03
 Lab Code: ACE SDG No.: Q2892
 Instrument ID: BNA_F Calibration Date/Time: 08/22/2025 10:29
 Lab File ID: BF143563.D Init. Calib. Date(s): 08/20/2025 08/20/2025
 EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:55 14:20
 GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.544	1.440		-6.7	
2,4-Dinitrotoluene	0.333	0.338		1.5	
Diethylphthalate	1.080	1.129		4.5	
4-Chlorophenyl-phenylether	0.594	0.562		-5.4	
Fluorene	1.188	1.122		-5.6	
4-Nitroaniline	0.248	0.264		6.5	
4,6-Dinitro-2-methylphenol	0.113	0.102		-9.7	
n-Nitrosodiphenylamine	0.644	0.599		-7.0	20.0
2,4,6-Tribromophenol	0.186	0.178		-4.3	
4-Bromophenyl-phenylether	0.239	0.222		-7.1	
Hexachlorobenzene	0.257	0.239		-7.0	
Atrazine	0.166	0.199		19.9	
Pentachlorophenol	0.118	0.108		-8.5	20.0
Phenanthrene	1.071	0.996		-7.0	
Anthracene	1.085	1.022		-5.8	
Carbazole	0.889	0.865		-2.7	
Di-n-butylphthalate	0.829	0.965		16.4	
Fluoranthene	0.961	0.923		-4.0	20.0
Pyrene	1.645	1.624		-1.3	
Terphenyl-d14	1.146	1.162		1.4	
Butylbenzylphthalate	0.449	0.488		8.7	
3,3-Dichlorobenzidine	0.369	0.335		-9.2	
Benzo (a) anthracene	1.288	1.184		-8.1	
Chrysene	1.157	1.118		-3.4	
Bis (2-ethylhexyl) phthalate	0.685	0.617		-9.9	
Di-n-octyl phthalate	1.268	1.115		-12.1	20.0
Benzo (b) fluoranthene	1.112	1.078		-3.1	
Benzo (k) fluoranthene	1.064	1.024		-3.8	
Benzo (a) pyrene	1.032	1.000		-3.1	20.0
Indeno (1,2,3-cd) pyrene	1.437	1.391		-3.2	
Dibenzo (a,h) anthracene	1.151	1.121		-2.6	
Benzo (g,h,i) perylene	1.145	1.103		-3.7	
1,2,4,5-Tetrachlorobenzene	0.572	0.527		-7.9	
1,4-Dioxane	0.531	0.503		-5.3	20.0
2,3,4,6-Tetrachlorophenol	0.286	0.274		-4.2	

All other compounds must meet a minimum RRF of 0.010.