

Surrogate Summary

SW-846

SDG No.: Q2892

Client: BAPS North Bergen Development Inc.

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169299BL	PB169299BL	2-Fluorophenol	150	125	84		30 (18)	130 (112)
		Phenol-d6	150	128	85		30 (15)	130 (107)
		Nitrobenzene-d5	100	95.9	96		30 (18)	130 (107)
		2-Fluorobiphenyl	100	86.3	86		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	133	89		30 (10)	130 (116)
		Terphenyl-d14	100	91.7	92		30 (10)	130 (105)
PB169299BS	PB169299BS	2-Fluorophenol	150	113	75		30 (18)	130 (112)
		Phenol-d6	150	114	76		30 (15)	130 (107)
		Nitrobenzene-d5	100	83.9	84		30 (18)	130 (107)
		2-Fluorobiphenyl	100	80.6	81		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	125	83		30 (10)	130 (116)
		Terphenyl-d14	100	91.7	92		30 (10)	130 (105)
Q2892-02	1-FLOOR-NORTH-EAST-CORNER	2-Fluorophenol	150	81.8	55		30 (18)	130 (112)
		Phenol-d6	150	80.6	54		30 (15)	130 (107)
		Nitrobenzene-d5	100	51.6	52		30 (18)	130 (107)
		2-Fluorobiphenyl	100	50.8	51		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	92.5	62		30 (10)	130 (116)
		Terphenyl-d14	100	50.6	51		30 (10)	130 (105)
Q2892-04	1-FLOOR-DRILL-B-BACK	2-Fluorophenol	150	74.5	50		30 (18)	130 (112)
		Phenol-d6	150	79.0	53		30 (15)	130 (107)
		Nitrobenzene-d5	100	58.9	59		30 (18)	130 (107)
		2-Fluorobiphenyl	100	55.3	55		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	76.7	51		30 (10)	130 (116)
		Terphenyl-d14	100	62.7	63		30 (10)	130 (105)
Q2892-06	BACK-1-FLOOR-CENTER-MIDDLE	2-Fluorophenol	150	81.4	54		30 (18)	130 (112)
		Phenol-d6	150	84.6	56		30 (15)	130 (107)
		Nitrobenzene-d5	100	59.3	59		30 (18)	130 (107)
		2-Fluorobiphenyl	100	61.0	61		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	82.9	55		30 (10)	130 (116)
		Terphenyl-d14	100	59.4	59		30 (10)	130 (105)
Q2892-08	3-FLOOR-NORTH-D	2-Fluorophenol	150	89.3	60		30 (18)	130 (112)
		Phenol-d6	150	91.9	61		30 (15)	130 (107)
		Nitrobenzene-d5	100	64.9	65		30 (18)	130 (107)
		2-Fluorobiphenyl	100	66.5	67		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	89.5	60		30 (10)	130 (116)
		Terphenyl-d14	100	65.9	66		30 (10)	130 (105)
Q2892-10	3-FLOOR-SOUTH-E	2-Fluorophenol	150	81.8	55		30 (18)	130 (112)
		Phenol-d6	150	85.0	57		30 (15)	130 (107)
		Nitrobenzene-d5	100	58.6	59		30 (18)	130 (107)
		2-Fluorobiphenyl	100	59.1	59		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	83.4	56		30 (10)	130 (116)
		Terphenyl-d14	100	65.1	65		30 (10)	130 (105)
Q2892-10MS	3-FLOOR-SOUTH-EMS	2-Fluorophenol	150	78.1	52		30 (18)	130 (112)
		Phenol-d6	150	79.0	53		30 (15)	130 (107)
		Nitrobenzene-d5	100	55.4	55		30 (18)	130 (107)
		2-Fluorobiphenyl	100	57.8	58		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	81.0	54		30 (10)	130 (116)
		Terphenyl-d14	100	59.8	60		30 (10)	130 (105)
Q2892-10MSD	3-FLOOR-SOUTH-EMSD	2-Fluorophenol	150	77.5	52		30 (18)	130 (112)
		Phenol-d6	150	77.5	52		30 (15)	130 (107)
		Nitrobenzene-d5	100	57.2	57		30 (18)	130 (107)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SW-846

SDG No.: Q2892

Client: BAPS North Bergen Development Inc.

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2892-10MSD	3-FLOOR-SOUTH-EMSD	2-Fluorobiphenyl	100	55.3	55		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	80.6	54		30 (10)	130 (116)
		Terphenyl-d14	100	59.9	60		30 (10)	130 (105)
Q2892-12	FRONT-1-FLOOR-F	2-Fluorophenol	150	66.5	44		30 (18)	130 (112)
		Phenol-d6	150	67.8	45		30 (15)	130 (107)
		Nitrobenzene-d5	100	48.5	49		30 (18)	130 (107)
		2-Fluorobiphenyl	100	49.6	50		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	68.6	46		30 (10)	130 (116)
		Terphenyl-d14	100	49.9	50		30 (10)	130 (105)