

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2892

Client ID: SSTDCCC040

Date Analyzed: 08/26/2025

Lab File ID: BP025579.D

Time Analyzed: 10:21

Instrument ID: BNA\_P

GC Column: ZB-GR ID: 0.25 (mm)

|                                | IS4 (PHN)<br>AREA # | RT #   | IS5 (CRY)<br>AREA # | RT #   | IS6 (PRY)<br>AREA # | RT #   |
|--------------------------------|---------------------|--------|---------------------|--------|---------------------|--------|
| 12 HOUR STD                    | 1130980             | 17.195 | 1110360             | 21.636 | 1292920             | 24.995 |
| UPPER LIMIT                    | 2261960             | 17.695 | 2220720             | 22.136 | 2585840             | 25.495 |
| LOWER LIMIT                    | 565490              | 16.695 | 555180              | 21.136 | 646460              | 24.495 |
| EPA SAMPLE NO.                 |                     |        |                     |        |                     |        |
| 01 1-FLOOR-NORTH-EAST-CORNER-1 | 1149710             | 17.20  | 1257630             | 21.64  | 1455180             | 25.01  |

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.