

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	Q2892	Analytical Method:	SW8260D
Client:	BAPS North Bergen Development Inc.	Datafile :	VY023165.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0820SBSD01	Dichlorodifluoromethane	20	21.4	ug/Kg	107	1		40 (64)	160 (136)	30 (20)
	Chloromethane	20	21.3	ug/Kg	106	2		40 (52)	160 (151)	30 (20)
	Vinyl chloride	20	21.7	ug/Kg	109	0		70 (56)	130 (148)	30 (20)
	Bromomethane	20	24.0	ug/Kg	120	2		40 (58)	160 (141)	30 (20)
	Chloroethane	20	21.8	ug/Kg	109	2		40 (69)	160 (130)	30 (20)
	Trichlorofluoromethane	20	20.8	ug/Kg	104	0		40 (69)	160 (134)	30 (20)
	1,1,2-Trichlorotrifluoroethane	20	21.4	ug/Kg	107	2		70 (81)	130 (123)	30 (20)
	1,1-Dichloroethene	20	20.3	ug/Kg	102	1		70 (79)	130 (121)	30 (20)
	Acetone	100	100	ug/Kg	100	10		40 (40)	160 (171)	30 (20)
	Carbon disulfide	20	20.3	ug/Kg	102	1		40 (59)	160 (130)	30 (20)
	Methyl tert-butyl Ether	20	18.5	ug/Kg	93	0		70 (77)	130 (129)	30 (20)
	Methyl Acetate	20	17.7	ug/Kg	89	5		70 (69)	130 (149)	30 (20)
	Methylene Chloride	20	19.6	ug/Kg	98	1		70 (72)	130 (131)	30 (20)
	trans-1,2-Dichloroethene	20	20.5	ug/Kg	103	1		70 (80)	130 (123)	30 (20)
	1,1-Dichloroethane	20	20.3	ug/Kg	102	0		70 (82)	130 (123)	30 (20)
	Cyclohexane	20	19.2	ug/Kg	96	1		70 (76)	130 (122)	30 (20)
	2-Butanone	100	98.0	ug/Kg	98	2		40 (69)	160 (131)	30 (20)
	Carbon Tetrachloride	20	20.3	ug/Kg	102	2		70 (76)	130 (129)	30 (20)
	cis-1,2-Dichloroethene	20	20.5	ug/Kg	103	1		70 (82)	130 (123)	30 (20)
	Bromochloromethane	20	20.1	ug/Kg	101	1		70 (80)	130 (127)	30 (20)
	Chloroform	20	20.9	ug/Kg	104	1		70 (82)	130 (125)	30 (20)
	1,1,1-Trichloroethane	20	20.7	ug/Kg	104	1		70 (80)	130 (126)	30 (20)
	Methylcyclohexane	20	19.3	ug/Kg	97	2		70 (77)	130 (123)	30 (20)
	Benzene	20	20.5	ug/Kg	103	1		70 (84)	130 (121)	30 (20)
	1,2-Dichloroethane	20	19.8	ug/Kg	99	0		70 (81)	130 (126)	30 (20)
	Trichloroethene	20	21.3	ug/Kg	106	0		70 (83)	130 (122)	30 (20)
	1,2-Dichloropropane	20	20.4	ug/Kg	102	0		70 (83)	130 (122)	30 (20)
	Bromodichloromethane	20	20.3	ug/Kg	102	0		70 (82)	130 (123)	30 (20)
	4-Methyl-2-Pentanone	100	88.2	ug/Kg	88	1		40 (70)	160 (135)	30 (20)
	Toluene	20	20.3	ug/Kg	102	1		70 (83)	130 (122)	30 (20)
	t-1,3-Dichloropropene	20	18.9	ug/Kg	95	2		70 (78)	130 (124)	30 (20)
	cis-1,3-Dichloropropene	20	19.5	ug/Kg	98	1		70 (81)	130 (122)	30 (20)
	1,1,2-Trichloroethane	20	19.8	ug/Kg	99	4		70 (82)	130 (125)	30 (20)
	2-Hexanone	100	90.9	ug/Kg	91	3		40 (66)	160 (138)	30 (20)
	Dibromochloromethane	20	20.2	ug/Kg	101	0		70 (79)	130 (125)	30 (20)
	1,2-Dibromoethane	20	19.9	ug/Kg	100	0		70 (80)	130 (125)	30 (20)
	Tetrachloroethene	20	22.9	ug/Kg	115	2		70 (83)	130 (125)	30 (20)
	Chlorobenzene	20	20.7	ug/Kg	104	2		70 (84)	130 (122)	30 (20)
	Ethyl Benzene	20	20.0	ug/Kg	100	1		70 (82)	130 (124)	30 (20)
	m/p-Xylenes	40	40.6	ug/Kg	102	1		70 (83)	130 (124)	30 (20)
	o-Xylene	20	19.9	ug/Kg	100	2		70 (83)	130 (123)	30 (20)
	Styrene	20	20.1	ug/Kg	101	0		70 (82)	130 (124)	30 (20)
	Bromoform	20	19.5	ug/Kg	98	0		70 (75)	130 (127)	30 (20)
	Isopropylbenzene	20	19.8	ug/Kg	99	2		70 (82)	130 (124)	30 (20)
	1,1,2,2-Tetrachloroethane	20	18.0	ug/Kg	90	2		70 (77)	130 (127)	30 (20)
	1,3-Dichlorobenzene	20	20.5	ug/Kg	103	1		70 (83)	130 (122)	30 (20)
	1,4-Dichlorobenzene	20	20.4	ug/Kg	102	2		70 (84)	130 (121)	30 (20)
	1,2-Dichlorobenzene	20	20.3	ug/Kg	102	3		70 (83)	130 (124)	30 (20)

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2892</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>BAPS North Bergen Development Inc.</u>	Datafile :	<u>VY023165.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0820SBSD01	1,2-Dibromo-3-Chloropropane	20	18.3	ug/Kg	92	4		40 (66)	160 (134)	30 (20)
	1,2,4-Trichlorobenzene	20	19.0	ug/Kg	95	4		70 (78)	130 (127)	30 (20)
	1,2,3-Trichlorobenzene	20	18.7	ug/Kg	94	2		70 (70)	130 (137)	30 (20)