

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: BAPS03
Lab Code: ACE SDG NO.: Q2892
Lab File ID: VY023162.D Date Analyzed: 08/20/2025
Instrument ID: MSVOA_Y Time Analyzed: 09:26
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	494314	7.71	753839	8.62	669405	11.41
UPPER LIMIT	988628	8.213	1507680	9.116	1338810	11.914
LOWER LIMIT	247157	7.213	376920	8.116	334703	10.914
EPA SAMPLE NO.						
1-FLOOR-NORTH-EAST-CORNER-BACK	490467	7.71	929671	8.62	806534	11.41
1-FLOOR-DRILL-B-BACK	500283	7.71	945353	8.62	831619	11.41
BACK-1-FLOOR-CENTER-MIDDLE-C	501346	7.71	951773	8.62	866589	11.41
3-FLOOR-NORTH-D	496959	7.71	948133	8.62	845587	11.41
3-FLOOR-SOUTH-E	494911	7.71	934233	8.62	840073	11.41
FRONT-1-FLOOR-F	486603	7.71	905782	8.62	805970	11.41
VY0820SBL01	578816	7.71	1083634	8.62	970907	11.42
VY0820SBS01	459555	7.71	730484	8.62	640306	11.41
VY0820SBSD01	458691	7.71	735089	8.62	636567	11.41

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.